

RECEIVED

OCT 06 2016

FDLE Alcohol Testing Program

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: HILLSBOROUGH CO SO
Time of Inspection: 17:38



Date of Inspection: 10/05/2016

Serial Number: 80-000830
Software: 8100.27

| Check or Test | YES | NO |
|---|-----|----|
| Date and/or Time Adjusted | | No |
| Diagnostic Check (Pre-Inspection): OK | Yes | |
| Alcohol Free Subject Test: 0.000 | Yes | |
| Mouth Alcohol Test: Slope Not Met | Yes | |
| Interferent Detect Test: Interferent Detect | Yes | |
| Diagnostic Check (Post-Inspection): OK | Yes | |


| Alcohol Free Test (g/210L) | 0.05g/210L Test (g/210L) Lot#:201507A Exp: 07/14/2017 | 0.08g/210L Test (g/210L) Lot#:201601F Exp: 01/26/2018 | 0.20g/210L Test (g/210L) Lot#:201604C Exp: 04/05/2018 | 0.08 g/210L Dry Gas Std Test (g/210L) Lot#:18015080A2 Exp: 08/05/2017 |
|----------------------------|---|---|---|---|
| 0.000 | 0.049 | 0.079 | 0.194 | 0.079 |
| 0.000 | 0.049 | 0.079 | 0.196 | 0.078 |
| 0.000 | 0.050 | 0.079 | 0.197 | 0.079 |

Number of Simulators Used: 5


Remarks:
AI C. KIRK 35602 present entire inspection. 


The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.


Signature and Printed Name

10/05/2016
Date

Reviewed
10-6-16
M. J. T. A. 
4:52