

Florida Department of Law Enforcement Alcohol Testing Program

JUN 10 2016

FDLE
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000 *BK*

Agency: LYNN HAVEN PD
Time of Inspection: 09:29

Date of Inspection: 06/09/2016

Serial Number: 80-000840
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: _____ Exp: _____	0.08g/210L Test (g/210L) Lot#: _____ Exp: _____	0.20g/210L Test (g/210L) Lot#: _____ Exp: _____	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: _____ Exp: _____

5071

Number of Simulators Used: _____

Remarks:

SKIPPED AI TO OPERATE INSTRUMENT; NOT A COMPLIANCE CHECK

The above instrument complies (*N/A*) does not comply (*PK*) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

PHILLIP W SUBER

Signature and Printed Name

06/09/2016
Date