

RECEIVED

Florida Department of Law Enforcement
Alcohol Testing Program

OCT 14 2016

FDLE
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: MONROE COUNTY SO
Time of Inspection: 05:16

Date of Inspection: 09/16/2016

Serial Number: 80-000867
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201507A Exp: 07/14/2017	0.08g/210L Test (g/210L) Lot#: 201601F Exp: 01/26/2018	0.20g/210L Test (g/210L) Lot#: 201505A Exp: 05/12/2017	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 181691B Exp: 02/13/2018
0.000	0.049	0.079	0.203	0.080
0.000	0.049	0.080	0.204	0.080
0.000	0.049	0.080	0.204	0.080

Number of Simulators Used: 5

Remarks:

A F / M A: Range Exceeded. **HAD RESIDUAL MOUTH ALCOHOL CAUSING THE RANGE EXCEEDED, THEN ~~REGISTERED~~ RE-RAN WITH NO ISSUE.**
J. Huff 8528/114

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

J. Huff 8528/114

Signature and Printed Name

JON B HUFF

09/16/2016
Date

RECEIVED

Florida Department of Law Enforcement

OCT 14 2016

Alcohol Testing Program

FDLE

Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: MONROE COUNTY S.O.
Time of Inspection: 04:35

Date of Inspection: 09/16/2016

Serial Number: 80-001254
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	


Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201507A Exp: 07/14/2017	0.08g/210L Test (g/210L) Lot#: 201601F Exp: 01/26/2018	0.20g/210L Test (g/210L) Lot#: 201505A Exp: 05/12/2017	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 1816818 Exp: 02/13/2018
0.000	0.051	0.080	0.201	0.080
0.000	0.050	0.081	0.203	0.079
0.000	0.051	0.080	0.203	0.079

Number of Simulators Used: 5

Remarks:
SPARE-GOING OUT FOR ANNUAL DEPT INSP

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

 8568/174

JON B HUFF

Signature and Printed Name

09/16/2016
Date

RECEIVED

Florida Department of Law Enforcement Alcohol Testing Program

OCT 14 2016

FDLE
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: MONROE COUNTY SO
Time of Inspection: 03:15

Date of Inspection: 09/15/2016

Serial Number: 80-000870
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201507A Exp: 07/14/2017	0.08g/210L Test (g/210L) Lot#: 201601F Exp: 01/26/2018	0.20g/210L Test (g/210L) Lot#: 201505A Exp: 05/12/2017	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 279870 Exp: 02/04/2019
0.000	0.050	0.081	0.205	0.030
0.000	0.050	0.082	0.206	0.080
0.000	0.051	0.082	0.204	0.080

Number of Simulators Used: 5

Remarks:
XW

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

8568/114

JON B HUFF

Signature and Printed Name

09/15/2016
Date

RECEIVED

Florida Department of Law Enforcement
Alcohol Testing Program

OCT 14 2016

FDLE
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: MONROE COUNTY SO
Time of Inspection: 02:36

Date of Inspection: 09/15/2016

Serial Number: 80-000868
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	


Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201507E Exp: 07/14/2017	0.08g/210L Test (g/210L) Lot#: 201601F Exp: 01/26/2018	0.20g/210L Test (g/210L) Lot#: 201505A Exp: 05/12/2017	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 279870 Exp: 02/04/2019
0.000	0.050	0.080	0.203	0.081
0.000	0.051	0.082	0.203	0.081
0.000	0.051	0.082	0.204	0.081

Number of Simulators Used: 5

Remarks:
KV-IN SERVICE

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

 8568/T14

JON B HUFF
Signature and Printed Name

09/15/2016
Date

RECEIVED

Florida Department of Law Enforcement OCT 14 2016

Alcohol Testing Program

FDLE

Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: MONROE COUNTY SO
Time of Inspection: 01:59

Date of Inspection: 09/15/2016

Serial Number: 80-000869
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201507A Exp: 07/14/2017	0.08g/210L Test (g/210L) Lot#:201601F Exp: 01/26/2018	0.20g/210L Test (g/210L) Lot#:201505A Exp: 05/12/2017	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:279870 Exp: 02/04/2019
0.000	0.051	0.081	0.204	0.082
0.000	0.051	0.082	0.204	0.082
0.000	0.051	0.082	0.205	0.082

Number of Simulators Used: 5

Remarks:

KV SPARE-GOING OUT FOR ANNUAL DEPT INSP

The above instrument complies (X) does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

8568/174

JON B HUFF

Signature and Printed Name

09/15/2016
Date

RECEIVED

Florida Department of Law Enforcement OCT 14 2016

Alcohol Testing Program

FDLE Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: MONROE COUNTY SO
Time of Inspection: 00:39

Date of Inspection: 09/16/2016

Serial Number: 80-006471
Software: 8100.27

Table with 3 columns: Check or Test, YES, NO. Rows include: Date and/or Time Adjusted, Diagnostic Check (Pre-Inspection): OK, Alcohol Free Subject Test: 0.000, Mouth Alcohol Test: Slope Not Met, Interferent Detect Test: Interferent Detect, Diagnostic Check (Post-Inspection): OK.

Table with 5 columns: Alcohol Free Test (g/210L), 0.05g/210L Test (g/210L), 0.08g/210L Test (g/210L), 0.20g/210L Test (g/210L), 0.08 g/210L Dry Gas Std Test (g/210L). Rows show test results for three different samples.

Number of Simulators Used: 5

Remarks: SPARE-GOING OUT FOR ANNUAL DEPT INSP

The above instrument complies (X) does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Handwritten signature and number 8568/114

JON B HUFF

Signature and Printed Name

09/16/2016
Date