

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: MIAMI PD
Time of Inspection: 12:55

Date of Inspection: 08/11/2016

Serial Number: 80-000874
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 000000 Exp: 08/14/2016	0.08g/210L Test (g/210L) Lot#: 000000 Exp: 08/14/2016	0.20g/210L Test (g/210L) Lot#: 000000 Exp: 08/14/2016	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 000000 Exp: 08/14/2016
0.000	0.052	0.083	0.205	0.079
0.000	0.052	0.083	0.206	0.079
0.000	0.052	0.083	0.206	0.079

Number of Simulators Used: 0

Remarks:
 was present

RECEIVED

SEP 23 2016

**FDLE
Alcohol Testing Program**

The above instrument complies (X) does not comply () with Chapter 110-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 110-8, FAC.

Signature and Printed Name

Date