

RECEIVED

# Florida Department of Law Enforcement Alcohol Testing Program

OCT 27 2016

FDLE  
Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: FWC  
Time of Inspection: 12:26

Date of Inspection: 10/15/2016

Serial Number: 80-000891  
Software: 8100.27

| Check or Test                               | YES | NO |
|---|-----|----|
| Date and/or Time Adjusted                   |     | No |
| Diagnostic Check (Pre-Inspection): OK       | Yes |    |
| Alcohol Free Subject Test: 0.000            |     | No |
| Mouth Alcohol Test: Slope Not Met           | Yes |    |
| Interferent Detect Test: Interferent Detect | Yes |    |
| Diagnostic Check (Post-Inspection): OK      | Yes |    |

| Alcohol Free Test (g/210L) | 0.05g/210L Test (g/210L)<br>Lot#:201507A<br>Exp: 07/14/2017 | 0.08g/210L Test (g/210L)<br>Lot#:201601F<br>Exp: 01/26/2018 | 0.20g/210L Test (g/210L)<br>Lot#:201605A<br>Exp: 05/12/2017 | 0.08 g/210L Dry Gas Std Test (g/210L)<br>Lot#:587124<br>Exp: 08/04/2019 |
|----------------------------|---|---|---|---|
| 0.000                      | 0.051   | 0.083   | 0.202   | 0.084   |
| 0.000                      | 0.052   | 0.083   | 0.204   | 0.084   |
| 0.000                      | 0.052   | 0.083   | 0.205   | 0.084   |

Number of Simulators Used: 5

Remarks: *Best Calibration for Alcohol Free Subject Test*

*JARROU T. HOLZAK*  
10/15/16

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

*JARROU T. HOLZAK*

JARROU T. HOLZAK  
Signature and Printed Name

10/15/2016  
Date