## Florida Department of Law Enforcement FEB 2 3 2016 Alcohol Testing Program Alcohol Testing Program Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency:	FFWCC	TAMPA	

Serial Number: 80-000902

Time of Inspection: 12:19

Date of Inspection: 02/22/2016

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: Exp:	0.08g/210L Test (g/210L) Lot#: Exp:	0.20g/210L Test (g/210L) Lot#: Exp:	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: Exp:

Numb	er of	Sir	nula	ator	s Used:					
Rema	rks:									
	BYPAS	SED	AΙ	TO	OPERATE	INSTRUMENT.	NOT	Α	COMPLIANCE	CHECK.

1/4 Dus

The above instrument complies ( $\lambda$ ) does not comply (	) with Chapter 11D-8, FAC.
I certify that I hold a valid Florida Department of La performed this inspection in accordance with the provision	
Yatubukall Signature and Pr	DANIELLE M BELL

02/22/2016