

# Florida Department of Law Enforcement Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: FFWCC SOUTH REGION  
Time of Inspection: 10:48

Date of Inspection: 09/22/2016

Serial Number: 80-000904  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201507A Exp: 07/14/2017	0.08g/210L Test (g/210L) Lot#:201601F Exp: 01/26/2018	0.20g/210L Test (g/210L) Lot#:201505A Exp: 05/12/2017	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:1757137 Exp: 10/15/2017
0.000	0.047	0.076	0.195	0.080
0.000	0.047	0.077	0.196	0.080
0.000	0.048	0.077	0.197	0.081

Number of Simulators Used: 5

Remarks:

A. F. / M. A.: Improper Sample.

*Note: Too much time lapse in mouth alcohol sample, had to repeat.*

*Chris Matt*

**RECEIVED**  
SEP 22 2016  
FDLE  
Alcohol Testing Program

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

*Chris Matt*

CHRIS J MATTSON

Signature and Printed Name

09/22/2016  
Date

RECEIVED

# Florida Department of Law Enforcement Alcohol Testing Program

JUL 13 2016

FDLE  
Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: FFWCC SOUTH REGION  
Time of Inspection: 11:35

Date of Inspection: 06/23/2016

Serial Number: 80-000904  
Software: 8100.27

Check or Test	YES	NO
<del>Date and/or Time Adjusted</del>		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201507A Exp: 07/14/2017	0.08g/210L Test (g/210L) Lot#:201502G Exp: 02/24/2017	0.20g/210L Test (g/210L) Lot#:201505A Exp: 05/12/2017	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:1757137 Exp: 10/15/2017
0.000	0.049	0.079	0.198	0.080
0.000	0.050	0.078	0.201	0.080
0.000	0.049	0.079	0.200	0.080

Number of Simulators Used: 5

Remarks:

A F / M A: Range Exceeded No Sample Provided.

*Range Exceeded -  
Reference not enough  
time lapse between  
sample & blowing into  
instrument  
Chris Mattson 7-12-16*

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

CHRIS J MATTSON

Signature and Printed Name

06/23/2016  
Date