

*del*

# Florida Department of Law Enforcement Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: **SANIBEL ISLAND PD**  
Time of Inspection: **01:35**

Date of Inspection: <sup>08/29/2016</sup>  
~~08/29/2000~~

Serial Number: **90-000937**  
Software: **8100.27**

Check or Test	YES	NO
Date and/or Time Adjusted	Yes	
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201603D Exp: 03/08/2018	0.08g/210L Test (g/210L) Lot#: 201601F Exp: 01/26/2018	0.20g/210L Test (g/210L) Lot#: 201604C Exp: 04/05/2018	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 32114080A2 Exp: 12/05/2016
0.000	0.049	0.080	0.194	0.081
0.000	0.050	0.081	0.196	0.081
0.000	0.050	0.081	0.197	0.081

Number of Simulators Used: 2

**Remarks:**

Time-Date changed.

Inspection date corrected

*MSGT. D L W*

**RECEIVED**

SEP 13 2016

**FDLE  
Alcohol Testing Program**

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

*MSGT. D L W* 201

DUSTIN L WORTH

Signature and Printed Name

08/29/2000  
Date

80-000937 *MS*

RECEIVED

# Florida Department of Law Enforcement Alcohol Testing Program

JUL 19 2016  
FDLE  
Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: SANIBEL ISLAND PD  
Time of Inspection: 21:59

Date of Inspection: 06/23/2016

Serial Number: 80-000937  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection) OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201507A Exp: 07/14/2017	0.08g/210L Test (g/210L) Lot#:201502G Exp: 02/24/2017	0.20g/210L Test (g/210L) Lot#:201505A Exp: 05/12/2017	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:32114080A2 Exp: 12/05/2016
/ 0.000	0.050	0.078	0.197	0.080
/ 0.000	0.050	0.080	0.197	0.080
/ 0.000	0.050	0.079	0.197	0.080

Number of Simulators Used: 2

Remarks:

00: Ambient Fail. - INSTRUMENT REPOSITIONED IN ROOM.

\* AMENDED

MSGT. *[Signature]* 20, 7/8/16

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

*MSGT. [Signature]* #201

DUSTIN L WORTH

Signature and Printed Name

06/23/2016  
Date