



# Florida Department of Law Enforcement Alcohol Testing Program

RECEIVED

JUL 20 2016

FDLE

Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: COLLIER COUNTY SO  
Time of Inspection: 11:03

Date of Inspection: 07/19/2016

Serial Number: 80-000944  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: _____ Exp: _____	0.08g/210L Test (g/210L) Lot#: _____ Exp: _____	0.20g/210L Test (g/210L) Lot#: _____ Exp: _____	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: _____ Exp: _____

*oom*

Number of Simulators Used: \_\_\_\_\_

Remarks:  
COMPLIANCE NOT DETERMINED, AI NOT CONDUCTED

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

*David E Reyes-Rivera*

DAVID E REYES-RIVERA

Signature and Printed Name

07/19/2016  
Date