

OCT 27 2016

Florida Department of Law Enforcement
Alcohol Testing Program

FDLE
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: CLEWISTON P.D.

Serial Number: 90-000933

Time of Inspection: 12:02

Date of Inspection: 10/20/2016

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201507A Exp: 07/14/2017	0.08g/210L Test (g/210L) Lot#: 201601F Exp: 01/26/2018	0.20g/210L Test (g/210L) Lot#: 201605A Exp: 05/12/2017	0.08 g/210L Dry Gas Std Test (g/210L) Lot #: 1737087 Exp: 09/12/2017
0.000	0.000 / 0.049	0.000 / 0.078	0.189 / 0.187	0.075
0.000	0.000 / 0.050	0.000 / 0.079	0.193 / 0.199	0.075
0.000	0.017 / 0.050	0.000 / 0.080	0.197 / 0.189	0.075

Number of Simulators Used: 1

Remarks:

05: Control Outside Tolerance/RESEALED AND RETESTED. 06: Control Outside Tolerance/FOUND CRACK IN JAR CHANGED. 20: Control Outside Tolerance/LOW TEMP REVEST.

Printer failed to print + unable to repair / new one ordered

The above instrument complies (X) does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

[Signature]
DARRIN A. MONAGHAN

Signature and Printed Name

10/20/2016

Date

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: CLEWISTON P.D.
Time of Inspection: 12:06

Date of Inspection: 09/29/2016

Serial Number: 80-000955
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: _____ Exp: _____	0.08g/210L Test (g/210L) Lot#: _____ Exp: _____	0.20g/210L Test (g/210L) Lot#: _____ Exp: _____	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: _____ Exp: _____

Number of Simulators Used: _____

Remarks:
BYPASSED AI TO OPERATE INSTRUMENT

*PHS
BK*

*N/A
COMPLIANCE NOT
DETERMINED*

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Shayla Platt

SHAYLA D PLATT

Signature and Printed Name

09/29/2016
Date