

RECEIVED

Florida Department of Law Enforcement Alcohol Testing Program

FEB 18 2016

FDLE

Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: ORANGE COUNTY S.O.
Time of Inspection: 10:25

Date of Inspection: 02/11/2016

Serial Number: 80-000963
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201411K Exp: 11/25/2016	0.08g/210L Test (g/210L) Lot#: 201502G Exp: 02/24/2017	0.20g/210L Test (g/210L) Lot#: 201505A Exp: 05/12/2017	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 05514080A1 Exp: 04/01/2016
0.000	0.050	0.080	0.203	0.089 / 0.082
0.000	0.051	0.080	0.204	0.088 / 0.083
0.000	0.050	0.081	0.205	0.089 / 0.082

Number of Simulators Used: 5

Remarks:

08: Control Outside Tolerance. CHANGED DGS LOT 05115080A5, EXP DATE 03/05/17. Repeated, 08 DGS Test, results were within the acceptable range. Instrument complies.

dm

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

KELLY M MELVILLE

Signature and Printed Name

02/11/2016
Date