

Florida Department of Law Enforcement Alcohol Testing Program

RECEIVED

OCT 06 2016

FDLE

Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: COCOA BEACH P.D.
Time of Inspection: 07:49

Date of Inspection: 10/06/2016

Serial Number: 80-000995
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted	<input checked="" type="checkbox"/>	<input type="checkbox"/>
Diagnostic Check (Pre-Inspection): OK	<input checked="" type="checkbox"/>	<input type="checkbox"/>
Alcohol Free Subject Test: 0.000	<input checked="" type="checkbox"/>	<input type="checkbox"/>
Mouth Alcohol Test: Slope Not Met	<input checked="" type="checkbox"/>	<input type="checkbox"/>
Interferent Detect Test: Interferent Detect	<input checked="" type="checkbox"/>	<input type="checkbox"/>
Diagnostic Check (Post-Inspection): OK	<input checked="" type="checkbox"/>	<input type="checkbox"/>

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: _____ Exp: _____	0.08g/210L Test (g/210L) Lot#: _____ Exp: _____	0.20g/210L Test (g/210L) Lot#: _____ Exp: _____	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: _____ Exp: _____

Number of Simulators Used: _____

Remarks:
YPASS TO OPERATE INSTRUMENT

** Not a compliance check*

SP ASK

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

JAKE L SHANAHAN

Signature and Printed Name

10/06/2016
Date