

999

RECEIVED

NOV 02 2016

Florida Department of Law Enforcement Alcohol Testing Program

FDLE
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: SEMINOLE COUNTY S.O.
Time of Inspection: 01:59

Date of Inspection: 10/29/2016

Serial Number: 80-000999
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	


Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201507A Exp: 07/14/2017	0.08g/210L Test (g/210L) Lot#:201601F Exp: 01/26/2018	0.20g/210L Test (g/210L) Lot#:201505A Exp: 05/12/2017	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:1787697 Exp: 01/08/2018
0.000	0.049	0.082	0.199	0.078
0.000	0.050	0.081	0.200	0.078
0.000	0.050	0.081	0.201	0.078

Number of Simulators Used: 5

Remarks:
Int Det:CONNECTION LOSE.

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



KEITH D BETHAM

Signature and Printed Name

10/29/2016
Date