

1000

RECEIVED

AUG 08 2016

# Florida Department of Law Enforcement Alcohol Testing Program

FDLE  
Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: COCOA P.D.  
Time of Inspection: 20:12

Date of Inspection: 08/02/2016

Serial Number: 80-001000  
Software: 8100.27

| Check or Test                               | YES | NO |
|---|-----|----|
| Date and/or Time Adjusted                   |     | No |
| Diagnostic Check (Pre-Inspection): OK       | Yes |    |
| Alcohol Free Subject Test: 0.000            | Yes |    |
| Mouth Alcohol Test: Slope Not Met           | Yes |    |
| Interferent Detect Test: Interferent Detect | Yes |    |
| Diagnostic Check (Post-Inspection): OK      | Yes |    |


| Alcohol Free Test (g/210L) | 0.05g/210L Test (g/210L)<br>Lot#:201507A<br>Exp: 07/14/2017 | 0.08g/210L Test (g/210L)<br>Lot#:201601F<br>Exp: 01/26/2018 | 0.20g/210L Test (g/210L)<br>Lot#:201505A<br>Exp: 05/12/2017 | 0.08 g/210L Dry Gas Std Test (g/210L)<br>Lot#:11315080A2<br>Exp: 05/05/2017 |
|----------------------------|---|---|---|---|
| 0.000                      | 0.050   | 0.079   | 0.201   | 0.079   |
| 0.000                      | 0.049   | 0.078   | 0.204   | 0.079   |
| 0.000                      | 0.051   | 0.080   | 0.203   | 0.079   |

Number of Simulators Used: 3

Remarks:

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



JAMES H FRAZIER

Signature and Printed Name

08/02/2016  
Date

# Florida Department of Law Enforcement

## Alcohol Testing Program *PJM ASK*

### AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: COCOA P.D.

Time of Inspection: 08:42

Date of Inspection: 03/07/2016

Serial Number: 80-001000

Software: 8100.27

| Check or Test                               | YES | NO |
|---|-----|----|
| Date and/or Time Adjusted                   |     | No |
| Diagnostic Check (Pre-Inspection): OK       |     | No |
| Alcohol Free Subject Test: 0.000            |     | No |
| Mouth Alcohol Test: Slope Not Met           |     | No |
| Interferent Detect Test: Interferent Detect |     | No |
| Diagnostic Check (Post-Inspection): OK      |     | No |

| Alcohol Free Test (g/210L) | 0.05g/210L Test (g/210L)<br>Lot#: Exp: | 0.08g/210L Test (g/210L)<br>Lot#: Exp: | 0.20g/210L Test (g/210L)<br>Lot#: Exp: | 0.08 g/210L Dry Gas Std Test (g/210L)<br>Lot#: Exp: |
|----------------------------|--|--|--|---|
|                            |  |  |  |   |
|                            |  |  |  |   |
|                            |  |  |  |   |

Number of Simulators Used: \_\_\_\_\_


Remarks:

BYPASS TO OPERATE INSTRUMENT Non-compliance: .

*\* Compliance Not Determined \**

The above instrument complies ( ) does not comply ( X ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

  
 Signature and Printed Name JAKE L SHANAHAN

03/07/2016  
 Date