

1006 SP

RECEIVED

Florida Department of Law Enforcement

NOV 23 2016

Alcohol Testing Program

FDLE
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: INDIALANTIC P.D.
Time of Inspection: 14:38

Date of Inspection: 10/23/2016

Serial Number: 80-001006
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201507A Exp: 07/14/2017	0.08g/210L Test (g/210L) Lot#:201601F Exp: 01/26/2018	0.20g/210L Test (g/210L) Lot#:201604C Exp: 04/05/2018	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:01316080A1 Exp: 02/05/2018
0.000	0.048	0.080	0.200	0.078
0.000	0.048	0.080	0.201	0.078
0.000	0.048	0.080	0.202	0.078


Number of Simulators Used: 3

Remarks:

Int Det: .
INT. DET TEST STARTED TO EARLY WITHOUT ADDING INT. INT. ADDED AND TEST REPEATED WITHOUT INCIDENT. 10/23/16

The above instrument complies (X) does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



RYAN M DOVEL

Signature and Printed Name

10/23/2016
Date

1006 SP

RECEIVED

Florida Department of Law Enforcement

NOV 23 2016

Alcohol Testing Program

FDLE
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: INDIALANTIC P.D.
Time of Inspection: 12:48

Date of Inspection: 10/23/2016

Serial Number: 80-001006
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: Exp:	0.08g/210L Test (g/210L) Lot#: Exp:	0.20g/210L Test (g/210L) Lot#: Exp:	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: Exp:
0.000				
0.000				
0.000				

Number of Simulators Used: 4

Remarks:

Int Det: . . Non-compliance: .

SIMULATOR MALFUNCTION, UNABLE TO MAKE PROPER SEAL, SIMULATOR REMOVED FROM USE: RM 10/23/16

The above instrument complies () does not comply (X) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

RYAN M. DOVEL

Signature and Printed Name

10/23/2016
Date