

RECEIVED

SEP 28 2016

1009
②

FDLE
Alcohol Testing Program

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: PALM BAY F.D.

Serial Number: 80-001009

Time of Inspection: 11:19

Date of Inspection: 09/25/2016

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201507A Exp: 07/14/2017	0.08g/210L Test (g/210L) Lot#:201601F Exp: 01/26/2018	0.20g/210L Test (g/210L) Lot#:201505A Exp: 05/12/2017	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:32114080A2 Exp: 12/05/2016
0.000	0.049	0.078	0.195	0.080
0.000	0.049	0.078	0.197	0.080
0.000	0.049	0.079	0.197	0.079

Number of Simulators Used: 5

Remarks:

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

EDWIN J LUTZ

Signature and Printed Name

09/25/2016
Date

RECEIVED

SEP 28 2016

1009
⑧

FDLE
Alcohol Testing Program

Florida Department of Law Enforcement
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: PALM BAY P.D.
Time of Inspection: 11:19

Date of Inspection: 09/25/2016

Serial Number: 80-001009
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201507A Exp: 07/14/2017	0.08g/210L Test (g/210L) Lot#:201601F Exp: 01/26/2018	0.20g/210L Test (g/210L) Lot#:201505A Exp: 05/12/2017	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:32114080A2 Exp: 12/05/2016
0.000	0.049	0.078	0.195	0.080
0.000	0.049	0.078	0.197	0.080
0.000	0.049	0.079	0.197	0.079

Number of Simulators Used: 5

Remarks:

The above instrument complies (X) does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

EDWIN J. LUTZ
Signature and Printed Name

09/25/2016
Date

1009

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: PALM BAY P.D.
 Time of Inspection: 10:20

Date of Inspection: 08/23/2016

Serial Number: 80-001009
 Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201507A Exp: 07/14/2017	0.08g/210L Test (g/210L) Lot#: 201601F Exp: 01/26/2018	0.20g/210L Test (g/210L) Lot#: 201505A Exp: 05/12/2017	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 32114080A2 Exp: 12/05/2016
0.000	0.049	0.078	0.195	0.079
0.000	0.049	0.078	0.196	0.079
0.000	0.049	0.078	0.198	0.079

Number of Simulators Used: 5

Remarks:

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

EDWIN J LUTZ

Signature and Printed Name

08/23/2016
 Date

D. Morgan # 46

1009
B

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: FLORIDA DEPT. OF LAW ENFORCEMENT Date of Inspection: 05/22/2016 Serial Number: 60-001600
Inspector: DAVID J. MORGAN Software: 3100.21

Agency Test	YES	NO
Calibration/Zero Adjusted		No
Pre-Inspection Work (Pre-Inspection): OK	Yes	
Pre-Inspection Subject Test: 0.00%	Yes	
Blank Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detected	Yes	
Post-Inspection Check (Post-Inspection): OK	Yes	

Alcohol Test	0.05g/210L Test (g/210L)	0.08g/210L Test (g/210L)	0.20g/210L Test (g/210L)	0.0% g/10L Dry Gas Std Test (g/210L)
Lot #	Lot#: 2011112	Lot #: 201020	Lot #: 210000	Lot #: 2011112
Exp. Date	Exp: 07/31/2017	Exp: 07/31/2017	Exp: 07/31/2018	Exp: 07/31/2017
Result	0.04%	0.07%	0.19%	0.00%
Result	0.04%	0.07%	0.19%	0.00%
Result	0.04%	0.07%	0.19%	0.00%

Signature of Inspector: _____
Date: _____

This instrument complies with the requirements of 21 CFR 101.11.1 with Feature ID# 8, #50
This report was prepared by David J. Morgan, Florida Department of Law Enforcement Agency Inspector. Performed and tested in accordance with the provisions of Chapter 110C, FSS.

Signature and Printed Name

Inspector
Date

Darin Morgan

Date: 05/22/2016

PEM Bay S.D.
Intoxilyzer - Alcohol Analyzer
Model 8500
05/22/2016
Software: 8100.27

PEM Bay S.D.
Intoxilyzer - Alcohol Analyzer
Model 8500
05/22/2016
Software: 8100.27

PEM Bay S.D.
Intoxilyzer - Alcohol Analyzer
Model 8500
05/22/2016
Software: 8100.27

PEM Bay S.D.
Intoxilyzer - Alcohol Analyzer
Model 8500
05/22/2016
Software: 8100.27

Test	9/21/16	Time
Air Blank	0.000	10:50
Control Test	0.194	10:51
Air Blank	0.000	10:51
Control Test	0.197	10:52
Air Blank	0.000	10:53
Control Test	0.196	10:53
Air Blank	0.000	10:54
Control Test Stats		
Average	0.1957	
Std Dev	0.0015	
Rel Std Dev(%)	0.7807	

[Signature]
Operator's Signature

Test	9/21/16	Time
Air Blank	0.000	10:48
Control Test	0.077	10:46
Air Blank	0.000	10:49
Control Test	0.078	10:49
Air Blank	0.000	10:50
Control Test	0.077	10:51
Air Blank	0.000	10:51
Control Test Stats		
Average	0.0773	
Std Dev	0.0006	
Rel Std Dev(%)	0.7466	

[Signature]
Operator's Signature

Test	9/21/16	Time
Air Blank	0.000	10:45
Control Test	0.047	10:47
Air Blank	0.000	10:47
Control Test	0.048	10:48
Air Blank	0.000	10:48
Control Test	0.048	10:49
Air Blank	0.000	10:51
Control Test Stats		
Average	0.0477	
Std Dev	0.0006	
Rel Std Dev(%)	1.2112	

[Signature]
Operator's Signature

PEM Bay S.D.
Intoxilyzer - Alcohol Analyzer
Model 8500
05/22/2016
Software: 8100.27

Test	9/21/16	Time
Air Blank	0.000	10:45
Control Test	0.000	10:46
Air Blank	0.000	10:47
Control Test	0.000	10:48
Air Blank	0.000	10:49
Control Test	0.000	10:50
Air Blank	0.000	10:51
Control Test Stats		
Average	0.0000	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

[Signature]
Operator's Signature

(INT. SOLUTION)