

RECEIVED

DEC 07 2016

# Florida Department of Law Enforcement Alcohol Testing Program

FDLE  
Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: BROWARD COUNTY SO  
Time of Inspection: 16:59

Date of Inspection: 11/23/2016

Serial Number: 80-001047  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201507A Exp: 07/14/2017	0.08g/210L Test (g/210L) Lot#:201502G Exp: 02/24/2017	0.20g/210L Test (g/210L) Lot#:201505A Exp: 05/12/2017	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:1789170 Exp: 01/14/2018
0.000	0.050	0.078	0.198	0.083
0.000	0.049	0.078	0.200	0.083
0.000	0.049	0.078	0.200	0.083

Number of Simulators Used: 5

Remarks:

STUART CAMPBELL WAS PRESENT AND OBSERVED ENTIRE INSPECTION.

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

FRANCA R. DICIZENZO  
Signature and Printed Name

11/23/2016  
Date

RECEIVED

SEP 27 2016

# Florida Department of Law Enforcement Alcohol Testing Program

FDLE  
Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: BROWARD COUNTY SO  
Time of Inspection: 22:10

Date of Inspection: 09/24/2016

Serial Number: 80-001047  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201507A Exp: 07/14/2017	0.08g/210L Test (g/210L) Lot#:201502G Exp: 02/24/2017	0.20g/210L Test (g/210L) Lot#:201505A Exp: 05/12/2017	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:1789170 Exp: 01/14/2018
0.000	0.048	0.078	0.200	0.080
0.000	0.049	0.079	0.203	0.080
0.000	0.049	0.078	0.203	0.080

Number of Simulators Used: 5

Remarks:  
A FRAZIER OBSERVED & Y WASHINGTON IN TRAINING

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

FRANCA R. DICIZENZO FRANCA R DICIZENZO  
Signature and Printed Name

09/24/2016  
Date

RECEIVED

Florida Department of Law Enforcement JUN 15 2016

Alcohol Testing Program

FDLE Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: BROWARD COUNTY SO
Time of Inspection: 09:20

Date of Inspection: 06/07/2016

Serial Number: 80-001047
Software: 8100.27

Table with 3 columns: Check or Test, YES, NO. Rows include Date and/or Time Adjusted, Diagnostic Check (Pre-Inspection): OK, Alcohol Free Subject Test: 0.000, Mouth Alcohol Test: Slope Not Met, Interferent Detect Test: Interferent Detect, Diagnostic Check (Post-Inspection): OK.

Table with 5 columns: Alcohol Free Test (g/210L), 0.05g/210L Test (g/210L), 0.08g/210L Test (g/210L), 0.20g/210L Test (g/210L), 0.08 g/210L Dry Gas Std Test (g/210L). Includes Lot# and Exp fields.

Number of Simulators Used: \_\_\_\_\_

Remarks: COMPLIANCE NOT DETERMINED AI NOT CONDUCTED

The above instrument complies ( X ) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Signature and Printed Name: DAVID E REYES-RIVERA

Date: 06/07/2016

RECEIVED

MAY 24 2016

# Florida Department of Law Enforcement Alcohol Testing Program

FDLE  
Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: BROWARD COUNTY SO  
Time of Inspection: 15:47

Date of Inspection: 04/25/2016

Serial Number: 80-001047  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201507A Exp: 07/14/2017	0.08g/210L Test (g/210L) Lot#:201502G Exp: 02/24/2017	0.20g/210L Test (g/210L) Lot#:N Exp: 05/12/2017	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:1789170 Exp: 01/14/2018
0.000	0.048	0.079	0.199	0.088 / 0.086
0.000	0.049	0.079	0.204	0.087 / 0.085
0.000	0.049	0.079	0.202	0.085 / 0.085

Number of Simulators Used: 5

Remarks:

08: Control Outside Tolerance, UNKNOWN REASON. A. FRAZIER OBS. \NSP.

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

*Franca R. Dicienzo*

FRANCA R DICIZENZO

Signature and Printed Name

04/25/2016  
Date

RECEIVED

MAY 24 2016

FDLE  
Alcohol Testing Program

BROWARD COUNTY SO  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-001247  
04/25/2016  
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	17:50
Control Test	0.000	17:50
Air Blank	0.000	17:51
Control Test	0.000	17:52
Air Blank	0.000	17:52
Control Test	0.000	17:53
Air Blank	0.000	17:53
Control Test	0.000	17:54
Air Blank	0.000	17:54
Control Test	0.000	17:55
Air Blank	0.000	17:55
Control Test Stats		
Average	0.0000	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

*Franca R. D. Cigo*  
Operator's Signature

*Int. not Detected  
on inspection, and  
also on Control Tests*

RECEIVED

MAY 24 2016

FDLE  
Alcohol Testing Program

BROWARD COUNTY SO  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-001249  
04/25/2016  
Software: 8100.27

-----  
WET CONTROL TEST

Test	g/210L	Time
Air Blank	0.000	17:46
Control Test	INT*	17:47
Air Blank	0.000	17:47

\*Interferent Detect

*Francis R. D'Agostino*  
Operator's Signature

\*Checking Interferent  
due to 80-001247  
not registering Interference  
when Int. Det. was  
introduced.

RECEIVED

MAR 02 2016

# Florida Department of Law Enforcement Alcohol Testing Program

FDLE  
Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: BROWARD COUNTY SO  
Time of Inspection: 20:25

Date of Inspection: 01/29/2016

Serial Number: 80-001047  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201507A Exp: 07/14/2017	0.08g/210L Test (g/210L) Lot#: 201502G Exp: 02/24/2017	0.20g/210L Test (g/210L) Lot#: 201505A Exp: 05/12/2017	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 1789170 Exp: 01/14/2018
0.000	0.052	0.080	0.203	0.083
0.000	0.052	0.080	0.204	0.082
0.000	0.051	0.080	0.204	0.081

Number of Simulators Used: 5

Remarks:

T SUN WAS PRESENT & OBSERVED ENTIRE INSPECTION

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

FRANCA R. DICIZENZO  
Signature and Printed Name

01/29/2016  
Date

RECEIVED

# Florida Department of Law Enforcement Alcohol Testing Program

JAN 21 2016

FDLE

Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: BROWARD COUNTY SO  
Time of Inspection: 07:51

Date of Inspection: 01/04/2016

Serial Number: 80-001047  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: _____ Exp: _____	0.08g/210L Test (g/210L) Lot#: _____ Exp: _____	0.20g/210L Test (g/210L) Lot#: _____ Exp: _____	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: _____ Exp: _____


Number of Simulators Used: \_\_\_\_\_

Remarks:

COMPLIANCE NOT DETERMINED, AI NOT CONDUCTED.

The above instrument complies ( X ) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



DAVID E REYES-RIVERA

Signature and Printed Name

01/04/2016  
Date