

RECEIVED

Florida Department of Law Enforcement

MAR 03 2016

Alcohol Testing Program

FDLE

Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: LAKE ALFRED PD

Serial Number: 80-001048

Time of Inspection: 13:41

Date of Inspection: 02/21/2016

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201507A Exp: 07/14/2017	0.08g/210L Test (g/210L) Lot#: 201502G Exp: 02/24/2017	0.20g/210L Test (g/210L) Lot#: 201505A Exp: 05/12/2017	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 05544080A1 Exp: 04/01/2016
0.000	0.048	0.076	0.197	0.081
0.000	0.049	0.077	0.198	0.080
0.000	0.049	0.077	0.200	0.080

Number of Simulators Used: 0

Remarks: OUT OF SERVICE/FDLE YR INSP

lower levels

The above instrument complies ( X ) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

CPL [Signature] #34

DANIEL L. ROYAL  
Signature and Printed Name

02/21/2016  
Date