

Florida Department of Law Enforcement Alcohol Testing Program

RECEIVED
JUL 22 2016
FDLE
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: ST PETERSBURG PD
Time of Inspection: 17:10

Date of Inspection: 07/14/2016

Serial Number: 80-001051
Software: 8100.27

| Check or Test | YES | NO |
|---|-----|----|
| Date and/or Time Adjusted | | No |
| Diagnostic Check (Pre-Inspection): OK | Yes | |
| Alcohol Free Subject Test: 0.000 | Yes | |
| Mouth Alcohol Test: Slope Not Met | Yes | |
| Interferent Detect Test: Interferent Detect | Yes | |
| Diagnostic Check (Post-Inspection): OK | | No |

| Alcohol Free Test (g/210L) | 0.05g/210L Test (g/210L) Lot#:201507A Exp: 07/14/2017 | 0.08g/210L Test (g/210L) Lot#:201601F Exp: 01/26/2018 | 0.20g/210L Test (g/210L) Lot#:201505A Exp: 05/12/2017 | 0.08 g/210L Dry Gas Std Test (g/210L) Lot#:28515080A2 Exp: 11/05/2017 |
|----------------------------|---|---|---|---|
| 0.000 | 0.049 | 0.079 | 0.203 | 0.000 / 0.045 |
| 0.000 | 0.050 | 0.080 | 0.204 | 0.000 / 0.079 |
| 0.000 | 0.050 | 0.080 | 0.204 | 0.000 / 0.078 |

Number of Simulators Used: 5

Remarks:

08: Control Outside Tolerance, STILL NOT CONNECT TUBE.

NEGLECTED TO CONNECT THE DRY GAS TUBE TWICE. INSPECTION FAILED DUE TO OPERATOR ERROR. SPOKE WITH DANIELLE BELL, SMD TO PERFORM INSPECTION CORRECTLY.

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

ROBBIE S ARKOVICH

Signature and Printed Name

07/14/2016
Date