

1054 SP

RECEIVED

# Florida Department of Law Enforcement Alcohol Testing Program

DEC 14 2016

FDLE  
Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: HIGHLANDS COUNTY SO  
Time of Inspection: 12:42

Date of Inspection: 12/11/2016

Serial Number: 80-001054  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	
	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201507A Exp: 07/14/2017	0.08g/210L Test (g/210L) Lot#: 201601F Exp: 01/26/2018	0.20g/210L Test (g/210L) Lot#: 201604C Exp: 04/05/2018	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 24515080A3 Exp: 10/05/2017
0.000	0.050	0.080	0.198	0.078
0.000	0.050	0.079	0.198	0.077
0.000	0.049	0.078	0.198	0.077

Number of Simulators Used: 5

Remarks:

**A TRUE COPY**  
**SUSAN BENTON, SHERIFF**  
**HIGHLANDS COUNTY, FLORIDA**  
 BY Shon W. Abeln 2184 D.S.

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Shon W. Abeln 2184

SHON W ABELN  
Signature and Printed Name

12/11/2016  
Date

1054 SR

RECEIVED

# Florida Department of Law Enforcement Alcohol Testing Program

DEC 01 2016  
FDLE  
Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: HIGHLANDS COUNTY SO  
Time of Inspection: 10:19

Date of Inspection: 11/27/2016

Serial Number: 80-001054  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201507A Exp: 07/14/2017	0.08g/210L Test (g/210L) Lot#:201601F Exp: 01/26/2018	0.20g/210L Test (g/210L) Lot#:201604C Exp: 04/05/2018	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:24515080A3 Exp: 10/05/2017
0.000	0.049	0.079	0.199	0.080
0.000	0.049	0.079	0.198	0.079
0.000	0.049	0.078	0.198	0.079

Number of Simulators Used: 5

Remarks:

A TRUE COPY  
SUSAN BENTON, SHERIFF  
HIGHLANDS COUNTY, FLORIDA  
BY Shon W. Abeln 2184 D.S.

The above instrument complies ( X ) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Shon W. Abeln 2184  
SHON W ABELN  
Signature and Printed Name

11/27/2016  
Date

RECEIVED

Florida Department of Law Enforcement OCT 28 2016  
Alcohol Testing Program

FDLE  
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: HIGHLANDS COUNTY SO  
Time of Inspection: 10:09

Date of Inspection: 10/23/2016

Serial Number: 80-001054  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201507A Exp: 07/14/2017	0.08g/210L Test (g/210L) Lot#: 201601F Exp: 01/26/2018	0.20g/210L Test (g/210L) Lot#: 201604C Exp: 04/05/2018	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 24515080A3 Exp: 10/05/2017
0.000	0.049	0.079	0.196	0.079
0.000	0.049	0.078	0.195	0.079
0.000	0.048	0.079	0.194	0.079

Number of Simulators Used: 5

Remarks:

A TRUE COPY  
SUSAN BENTON, SHERIFF  
HIGHLANDS COUNTY, FLORIDA  
BY Shon W. Abeln 2184 D.S.

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Shon W. Abeln 2184

Signature and Printed Name

SHON W ABELN

SP

10/23/2016  
Date

RECEIVED

Florida Department of Law Enforcement SEP 28 2016  
Alcohol Testing Program

FDLE  
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: HIGHLANDS COUNTY SO  
Time of Inspection: 07:45

Date of Inspection: 09/25/2016

Serial Number: 80-001054  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201507A Exp: 07/14/2017	0.08g/210L Test (g/210L) Lot#:201601F Exp: 01/26/2018	0.20g/210L Test (g/210L) Lot#:201604C Exp: 04/05/2018	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:24515080A3 Exp: 10/05/2017
0.000	0.049	0.078	0.197	0.079
0.000	0.048	0.079	0.196	0.079
0.000	0.049	0.078	0.195	0.078

Number of Simulators Used: 5

Remarks:

A TRUE COPY  
SUSAN BENTON, SHERIFF  
HIGHLANDS COUNTY, FLORIDA  
BY Shon W. Abeln D.S.

The above instrument complies ( X ) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

SHON W ABELN

Signature and Printed Name

09/25/2016  
Date

RECEIVED

# Florida Department of Law Enforcement Alcohol Testing Program

AUG 22 2016

FDLE  
Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: HIGHLANDS COUNTY SO  
Time of Inspection: 13:45

Date of Inspection: 08/19/2016

Serial Number: 80-001054  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201507A Exp: 07/14/2017	0.08g/210L Test (g/210L) Lot#:201502G Exp: 02/24/2017	0.20g/210L Test (g/210L) Lot#:201505A Exp: 05/12/2017	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:24515080A3 Exp: 10/05/2017
0.000	0.049	0.077	0.197	0.079
0.000	0.049	0.077	0.198	0.079
0.000	0.049	0.077	0.197	0.079

Number of Simulators Used: 5

Remarks:

A TRUE COPY  
SUSAN DENTON, SHERIFF  
HIGHLANDS COUNTY, FLORIDA  
BY  D.S.

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

GREGORY S GLASCO

Signature and Printed Name

08/19/2016  
Date

RECEIVED

Florida Department of Law Enforcement JUN 21 2016

Alcohol Testing Program

FDLE Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: HIGHLANDS COUNTY SO
Time of Inspection: 14:16

Date of Inspection: 06/17/2016

Serial Number: 80-001054
Software: 8100.27

Table with 3 columns: Check or Test, YES, NO. Rows include Date and/or Time Adjusted, Diagnostic Check (Pre-Inspection): OK, Alcohol Free Subject Test: 0.000, Mouth Alcohol Test: Slope Not Met, Interferent Detect Test: Interferent Detect, Diagnostic Check (Post-Inspection): OK.

Table with 5 columns: Alcohol Free Test (g/210L), 0.05g/210L Test (g/210L), 0.08g/210L Test (g/210L), 0.20g/210L Test (g/210L), 0.08 g/210L Dry Gas Std Test (g/210L). Rows show test results for three different samples.

Number of Simulators Used: 5

Remarks:

A TRUE COPY
SUSAN BENTON, SHERIFF
HIGHLANDS COUNTY, FLORIDA
BY [Signature] D.S.

The above instrument complies ( X ) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

GREGORY S GLASCO

Signature and Printed Name

06/17/2016
Date

# Florida Department of Law Enforcement Alcohol Testing Program

MAY 25 2016

FDLE  
Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: HIGHLANDS COUNTY SO  
Time of Inspection: 08:36

Date of Inspection: 05/21/2016

Serial Number: 80-001054  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201507A Exp: 07/14/2017	0.08g/210L Test (g/210L) Lot#:201502G Exp: 02/24/2017	0.20g/210L Test (g/210L) Lot#:201505A Exp: 05/12/2017	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:24515080A3 Exp: 10/05/2017
0.000	0.049	0.078	0.197	0.080
0.000	0.050	0.077	0.198	0.079
0.000	0.048	0.077	0.198	0.079

Number of Simulators Used: 5

Remarks:

**A TRUE COPY**  
**SUSAN BENTON, SHERIFF**  
**HIGHLANDS COUNTY, FLORIDA**  
BY Shon W. Abeln 2184 D.S.

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Shon W. Abeln 2184

SHON W ABELN  
Signature and Printed Name

05/21/2016  
Date

RECEIVED

# Florida Department of Law Enforcement

## Alcohol Testing Program

MAY 02 2016

FDLE

Alcohol Testing Program

### AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: HIGHLANDS COUNTY SO  
Time of Inspection: 11:00

Date of Inspection: 04/29/2016


Serial Number: 80-001054  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201507A Exp: 07/14/2017	0.08g/210L Test (g/210L) Lot#:201502G Exp: 02/24/2017	0.20g/210L Test (g/210L) Lot#:201505A Exp: 05/12/2017	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:24515080A3 Exp: 10/05/2017
0.000	0.047	0.076	0.197	0.078
0.000	0.048	0.077	0.198	0.078
0.000	0.047	0.078	0.197	0.078

Number of Simulators Used: 5

Remarks:

A TRUE COPY  
SUSAN BENTON, SHERIFF  
HIGHLANDS COUNTY, FLORIDA  
BY  D.S.

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

GREGORY S GLASCO

Signature and Printed Name

04/29/2016  
Date



RECEIVED

Florida Department of Law Enforcement  
Alcohol Testing Program

MAR 21 2016

FDLE  
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: HIGHLANDS COUNTY SO  
Time of Inspection: 14:35

Date of Inspection: 03/18/2016

Serial Number: 80-001054  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201507A Exp: 07/14/2017	0.08g/210L Test (g/210L) Lot#: 201502G Exp: 02/24/2017	0.20g/210L Test (g/210L) Lot#: 201505A Exp: 05/12/2017	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 24515080A3 Exp: 10/05/2017
0.000	0.049	0.077	0.198	0.079
0.000	0.048	0.077	0.199	0.079
0.000	0.048	0.077	0.198	0.079


Number of Simulators Used: 5

Remarks:

A TRUE COPY  
SUSAN BENTON, SHERIFF  
HIGHLANDS COUNTY, FLORIDA  
BY  D.S.

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

  
Signature and Printed Name GREGORY S GLASCO

03/18/2016  
Date

# Florida Department of Law Enforcement Alcohol Testing Program

RECEIVED

FEB 22 2016

FDLE  
Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: HIGHLANDS COUNTY SO  
Time of Inspection: 13:25

Date of Inspection: 02/19/2016

Serial Number: 80-001054  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201411K Exp: 11/25/2016	0.08g/210L Test (g/210L) Lot#:201502G Exp: 02/24/2017	0.20g/210L Test (g/210L) Lot#:201408B Exp: 08/12/2016	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:24515080A3 Exp: 10/05/2017
0.000	0.047 / 0.046	0.075	0.192	0.079
0.000	0.047 / 0.048	0.076	0.195	0.079
0.000	/ 0.048	0.077	0.198	0.079

Number of Simulators Used: 5

Remarks:

05: RFI Detect.

*HAD my cell phone to close - retested*

A TRUE COPY  
SUSAN BENTON, SHERIFF  
HIGHLANDS COUNTY, FLORIDA  
BY  D.S.

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

GREGORY S GLASCO

Signature and Printed Name

02/19/2016  
Date

Florida Department of Law Enforcement JAN 19 2016

Alcohol Testing Program

FDLE Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: HIGHLANDS COUNTY SO  
Time of Inspection: 14:10

Date of Inspection: 01/15/2016

Serial Number: 80-001054  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201411K Exp: 11/25/2016	0.08g/210L Test (g/210L) Lot#:201502G Exp: 02/24/2017	0.20g/210L Test (g/210L) Lot#:201408B Exp: 08/12/2016	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:24515080A3 Exp: 10/05/2017
0.000	0.048	0.078	0.197	0.078
0.000	0.048	0.078	0.199	0.078
0.000	0.048	0.078	0.200	0.078


Number of Simulators Used: 5

Remarks:

A TRUE COPY  
SUSAN BENSON, SHERIFF  
HIGHLANDS COUNTY, FLORIDA  
BY  D.S.

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

 \_\_\_\_\_  
Signature and Printed Name GREGORY S GLASCO

01/15/2016  
Date