

Florida Department of Law Enforcement Alcohol Testing Program

AUG 10 2016

FDLE
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: DAVIE PD

Time of Inspection: 14:30

Date of Inspection: 07/22/2016

Serial Number: 80-001056

Software: 8100.27.00

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201507A Exp: 07/14/2017	0.08g/210L Test (g/210L) Lot#:201502G Exp: 02/24/2017	0.20g/210L Test (g/210L) Lot#:201505A Exp: 05/12/2017	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG429701 Exp: 10/24/2016
/ 0.000	0.048	0.081	0.205	0.079
/ 0.000	0.049	0.081	0.206	0.079
/ 0.000	0.049	0.081	0.206	0.079

Number of Simulators Used: 1

Remarks:

00: Ambient Fail.

UNKNOWN REASON FOR AMBIENT FAIL ON FIRST
DIAGNOSTIC CHECK.
SAT M. 2 #593

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

MARK B LEONE

Signature and Printed Name

07/22/2016
Date