

THOMSON

RECEIVED

Florida Department of Law Enforcement Alcohol Testing Program

JUN 09 2016

FDLE

AGENCY INSPECTION REPORT - INTOXILYZER 8000 Alcohol Testing Program

Agency: JACKSONVILLE SO
Time of Inspection: 07:28

Date of Inspection: 05/13/2016

Serial Number: 80-001069
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201411K Exp: 11/25/2016	0.08g/210L Test (g/210L) Lot#:201502G Exp: 02/24/2017	0.20g/210L Test (g/210L) Lot#:201508B Exp: 08/12/2016	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:24515080A2 Exp: 10/05/2017
0.000	0.048	0.077	0.200	0.080
0.000	0.048	0.077	0.199	0.080
0.000	0.047	0.077	0.200	0.080

Number of Simulators Used: 5

Remarks:

*LOT NUMBER ON .020 SHOULD BE 201408B
NOT 201508B. RDT 7326*

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Robert D Thomson 7326
ROBERT D THOMASON
Signature and Printed Name

05/13/2016
Date

Amended

Florida Department of Law Enforcement Alcohol Testing Program

APR 29 2016
FDLE
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: JACKSONVILLE SO
Time of Inspection: 08:50

Date of Inspection: 04/28/2016

Serial Number: 80-001069
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201411K Exp: 11/25/2016	0.08g/210L Test (g/210L) Lot#:201502G Exp: 02/24/2017	0.20g/210L Test (g/210L) Lot#:201408B Exp: 08/12/2016	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:24515080A2 Exp: 10/05/2017
0.000	0.047 / 0.047	0.078	0.199	0.080
0.000	0.047 / 0.047	0.078	0.199	0.079
0.000	/ 0.047	0.077	0.199	0.079

Number of Simulators Used: 5

Remarks:
05: RFI Detect.

*NOTES: UNK. RFI CHECKED ROOM RETEST O.K.
PS17326*

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Robert D Thomason

Signature and Printed Name ROBERT D THOMASON

04/28/2016
Date