

RECEIVED

Florida Department of Law Enforcement  
 Alcohol Testing Program

AUG 30 2016

FDLE

Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: HARDEE COUNTY SO  
 Time of Inspection: 14:05

Date of Inspection: 08/29/2016

Serial Number: 80-001077  
 Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201507A Exp: 07/14/2017	0.08g/210L Test (g/210L) Lot#:201601F Exp: 01/26/2018	0.20g/210L Test (g/210L) Lot#:201505A Exp: 05/12/2017	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:1720515 Exp: 07/31/2017
0.000	0.046	0.078	0.198	0.074 / 0.076
0.000	0.047	0.078	0.199	0.074 / 0.075
0.000	0.047	0.078	0.199	0.074 / 0.075

Number of Simulators Used: 5

Remarks:

08: Control Outside Tolerance. Dry Gas std test. Nothing changed between First test and second test. Just reran test. After completion of test ran stability test and determined that the dry gas was the issue.

*Sgt Gary W. Cartwright* 8/29/16

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

*Gary W. Cartwright*

GARY W. CARTWRIGHT JR

Signature and Printed Name

08/29/2016  
 Date

80-001077 *DWB*

RECEIVED

# Florida Department of Law Enforcement Alcohol Testing Program

FEB 24 2015

FDLE

Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: HARDEE COUNTY SO  
Time of Inspection: 12:43

Date of Inspection: 02/23/2016

Serial Number: 80-001077  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201507A Exp: 07/14/2017	0.08g/210L Test (g/210L) Lot#: 201502G Exp: 02/24/2017	0.20g/210L Test (g/210L) Lot#: 201408B Exp: 08/12/2016	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 1720515 Exp: 07/31/2017
/ 0.000	0.047	0.081	0.197	0.078
/ 0.000	0.048	0.081	0.199	0.079
/ 0.000	0.048	0.081	0.199	0.078

Number of Simulators Used: 5

Remarks: Alcohol Free test

00: Ambient fail. Due to I did not have the simulator attached and started the test. I ~~did not attach~~ attached the simulator after I started the test.

Taken out of service for yearly inspection

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

GARY W CARTWRIGHT JR

Signature and Printed Name

02/23/2016  
Date