

RECEIVED

Florida Department of Law Enforcement Alcohol Testing Program

JAN 20 2016

FDLE
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: LARGO PD
Time of Inspection: 13:10

Date of Inspection: 01/20/2016

Serial Number: 80-001079
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201411K Exp: 11/25/2016	0.08g/210L Test (g/210L) Lot#: 201502G Exp: 02/24/2017	0.20g/210L Test (g/210L) Lot#: 201408B Exp: 08/12/2016	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 05115080A5 Exp: 03/05/2017
0.000	0.048	0.077	0.194	0.081
0.000	0.049	0.078	0.200	0.081
0.000	0.048	0.078	0.200	0.081

Number of Simulators Used: 5

Remarks:

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

R. Cook #460

ROBERT C. COOK

Signature and Printed Name

01/20/2016
Date

Florida Department of Law Enforcement Alcohol Testing Program

JAN 20 2016

FDLE
Alcohol Testing Program

AGENCY INSPECTION REPORT -- INTOXILYZER 8000

Agency: LARGO PD
Time of Inspection: 11:42

Date of Inspection: 01/20/2016

Serial Number: 80-001079
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201411K Exp: 11/25/2016	0.08g/210L Test (g/210L) Lot#: 201402U Exp: 02/24/2017	0.20g/210L Test (g/210L) Lot#: 201402H Exp: 08/12/2016	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: Exp:
0.000	0.047	0.076	0.189 / 0.186	
0.000	0.047	0.076	0.193 / 0.185	
0.000	0.047	0.075	0.191 / 0.183	

Number of Simulators Used: 5

Remarks:

20: Control Outside Tolerance, Control Outside Tolerance. Non-compliance: 20 SOLUTION FINDER NOT ACCRGE EITHER APP.

- Instr. not in compl., taken out of serv.
- Contacted Ms. Bell @ FDLE.
- Conducted stability test (3x) on 20 solution
- 20 solution was now w/in tolerance (.190-.210)
- 2nd inspection conducted w/in compliance
- Instr. placed back in service.

The above instrument complies () does not comply (X) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

R. Cook #460

ROBERT S. COOK

Signature and Printed Name

01/20/2016
Date