

RECEIVED

JUL 05 2016

FDLE

Alcohol Testing Program

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: Polk County Sheriff's Office

Serial Number: 80-001080

Time of Inspection: 1:10

Date of Inspection: 7/5/16

Software: 8100.27

CHECK OR TEST	YES	NO
Date and/or Time Adjusted		
Diagnostic Check (Pre-Inspection): OK		
Alcohol Free Subject Test: 0.000		
Mouth Alcohol Test: Slope Not Met		
Interferent Detect Test: Interferent Detect		
Diagnostic Check (Post-Inspection): OK		


Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: _____ Exp: _____	0.08g/210L Test (g/210L) Lot#: _____ Exp: _____	0.20g/210L Test (g/210L) Lot#: _____ Exp: _____	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: _____ Exp: _____

Number of Simulators Used: _____

Remarks: While attempting to conduct a Monthly Inspection on the instrument, I continued to get a RFI Detect error and was unable to complete the inspection. The instrument was removed from evidential use. The instrument is being sent to FDLE for inspection and/or repair.

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



Barbie R Wilson, 6493

Signature and Printed Name

7/5/16

Date

Florida Department of Law Enforcement RECEIVED

Alcohol Testing Program

JAN 05 2016

AGENCY INSPECTION REPORT - INTOXILYZER 8000 FDLE

Alcohol Testing Program

Agency: POLK COUNTY SO
 Time of Inspection: 12:15

Date of Inspection: 01/05/2016

Serial Number: 80-001080
 Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201507A Exp: 07/14/2017	0.08g/210L Test (g/210L) Lot#: 201502G Exp: 02/24/2017	0.20g/210L Test (g/210L) <i>SA</i> Lot#: 201502G Exp: 02/24/2017 <i>5112</i>	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 18015080A2 Exp: 08/05/2017
0.000	0.049	0.079	0.202	0.077
0.000	0.049	0.079	0.203	0.077
0.000	0.049	0.080	0.204	0.078

Number of Simulators Used: 5

Remarks:

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

[Signature] 6493 BARBIE R WILSON
 Signature and Printed Name

01/05/2016
 Date