

RECEIVED  
DEC 01 2016

FDLE  
Alcohol Testing Program

# Florida Department of Law Enforcement Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: PASCO COUNTY SO  
Time of Inspection: 11:55

Date of Inspection: 12/01/2016

Serial Number: 80-001081  
Software: 8100.27

| Check or Test                               | YES | NO |
|---|-----|----|
| Date and/or Time Adjusted                   |     | No |
| Diagnostic Check (Pre-Inspection): OK       | Yes |    |
| Alcohol Free Subject Test: 0.000            | Yes |    |
| Mouth Alcohol Test: Slope Not Met           | Yes |    |
| Interferent Detect Test: Interferent Detect | Yes |    |
| Diagnostic Check (Post-Inspection): OK      | Yes |    |

| Alcohol Free Test (g/210L) | 0.05g/210L Test (g/210L)<br>Lot#:201507A<br>Exp: 07/14/2017 | 0.08g/210L Test (g/210L)<br>Lot#:201601F<br>Exp: 01/26/2018 | 0.20g/210L Test (g/210L)<br>Lot#:201505A<br>Exp: 05/12/2017 | 0.08 g/210L Dry Gas Std Test (g/210L)<br>Lot#:03415080A1<br>Exp: 03/05/2017 |
|----------------------------|---|---|---|---|
| 0.000                      | 0.049   | 0.077   | 0.198   | 0.079   |
| 0.000                      | 0.049   | 0.077   | 0.199   | 0.078   |
| 0.000                      | 0.049   | 0.078   | 0.200   | 0.078   |

Number of Simulators Used: 5

Remarks:

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Ricky A Frisco 3184 RICKY A FRISCO  
Signature and Printed Name

12/01/2016  
Date

80-001081 @DS

RECEIVED

NOV 03 2016

# Florida Department of Law Enforcement Alcohol Testing Program

FDLE  
Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: PASCO COUNTY SO  
Time of Inspection: 09:13

Date of Inspection: 11/03/2016

Serial Number: 80-001081  
Software: 8100.27

| Check or Test                               | YES | NO |
|---|-----|----|
| Date and/or Time Adjusted                   |     | No |
| Diagnostic Check (Pre-Inspection): OK       | Yes |    |
| Alcohol Free Subject Test: 0.000            | Yes |    |
| Mouth Alcohol Test: Slope Not Met           | Yes |    |
| Interferent Detect Test: Interferent Detect | Yes |    |
| Diagnostic Check (Post-Inspection): OK      | Yes |    |

| Alcohol Free Test (g/210L) | 0.05g/210L Test (g/210L)<br>Lot#:201507A<br>Exp: 07/14/2017 | 0.08g/210L Test (g/210L)<br>Lot#:201601F<br>Exp: 01/26/2018 | 0.20g/210L Test (g/210L)<br>Lot#:201505A<br>Exp: 05/12/2017 | 0.08 g/210L Dry Gas Std Test (g/210L)<br>Lot#:03415080A1<br>Exp: 03/05/2017 |
|----------------------------|---|---|---|---|
| 0.000                      | 0.048   | 0.077   | 0.199   | 0.077   |
| 0.000                      | 0.048   | 0.078   | 0.200   | 0.077   |
| 0.000                      | 0.049   | 0.079   | 0.201   | 0.078   |

Number of Simulators Used: 5

Remarks:

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Ricky A. Frisco  
3184  
Signature and Printed Name

11/03/2016  
Date

# Florida Department of Law Enforcement Alcohol Testing Program

**RECEIVED**  
OCT 14 2016  
FDLE  
Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: PASCO COUNTY SO  
Time of Inspection: 08:34

Date of Inspection: 10/14/2016

Serial Number: 80-001081  
Software: 8100.27

| Check or Test                               | YES | NO |
|---|-----|----|
| Date and/or Time Adjusted                   |     | No |
| Diagnostic Check (Pre-Inspection): OK       | Yes |    |
| Alcohol Free Subject Test: 0.000            | Yes |    |
| Mouth Alcohol Test: Slope Not Met           | Yes |    |
| Interferent Detect Test: Interferent Detect | Yes |    |
| Diagnostic Check (Post-Inspection): OK      | Yes |    |

| Alcohol Free Test (g/210L) | 0.05g/210L Test (g/210L)<br>Lot#:201507A<br>Exp: 07/14/2017 | 0.08g/210L Test (g/210L)<br>Lot#:201601F<br>Exp: 01/26/2018 | 0.20g/210L Test (g/210L)<br>Lot#:201505A<br>Exp: 05/12/2017 | 0.08 g/210L Dry Gas Std Test (g/210L)<br>Lot#:03415080A1<br>Exp: 03/05/2017 |
|----------------------------|---|---|---|---|
| 0.000                      | 0.049   | 0.077   | 0.198   | 0.077   |
| 0.000                      | 0.049   | 0.078   | 0.199   | 0.077   |
| 0.000                      | 0.050   | 0.077   | 0.200   | 0.078   |

Number of Simulators Used: 5

Remarks:

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

*Ricky A. Frisco*      RICKY A FRISCO  
3184      Signature and Printed Name

10/14/2016  
Date

# Florida Department of Law Enforcement Alcohol Testing Program

**RECEIVED**

SEP 15 2016

FDLE  
Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: PASCO COUNTY SO  
Time of Inspection: 10:21

Date of Inspection: 09/15/2016

Serial Number: 80-001081  
Software: 8100.27

| Check or Test                               | YES | NO |
|---|-----|----|
| Date and/or Time Adjusted                   |     | No |
| Diagnostic Check (Pre-Inspection): OK       | Yes |    |
| Alcohol Free Subject Test: 0.000            | Yes |    |
| Mouth Alcohol Test: Slope Not Met           | Yes |    |
| Interferent Detect Test: Interferent Detect | Yes |    |
| Diagnostic Check (Post-Inspection): OK      | Yes |    |


| Alcohol Free Test (g/210L) | 0.05g/210L Test (g/210L)<br>Lot#:201507A<br>Exp: 07/14/2017 | 0.08g/210L Test (g/210L)<br>Lot#:201601F<br>Exp: 01/26/2018 | 0.20g/210L Test (g/210L)<br>Lot#:201505A<br>Exp: 05/12/2017 | 0.08 g/210L Dry Gas Std Test (g/210L)<br>Lot#:03415080A1<br>Exp: 03/05/2017 |
|----------------------------|---|---|---|---|
| 0.000                      | 0.048   | 0.077   | 0.197   | 0.078   |
| 0.000                      | 0.049   | 0.077   | 0.198   | 0.078   |
| 0.000                      | 0.049   | 0.079   | 0.200   | 0.078   |

Number of Simulators Used: 5

Remarks:

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



RICKY A PRISCO  
Signature and Printed Name

09/15/2016  
Date

RECEIVED

AUG 19 2016

FDLE  
Alcohol Testing Program

# Florida Department of Law Enforcement Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: PASCO COUNTY SO  
Time of Inspection: 10:54

Date of Inspection: 08/19/2016

Serial Number: 80-001081  
Software: 8100.27

| Check or Test                               | YES | NO |
|---|-----|----|
| Date and/or Time Adjusted                   |     | No |
| Diagnostic Check (Pre-Inspection): OK       | Yes |    |
| Alcohol Free Subject Test: 0.000            | Yes |    |
| Mouth Alcohol Test: Slope Not Met           | Yes |    |
| Interferent Detect Test: Interferent Detect | Yes |    |
| Diagnostic Check (Post-Inspection): OK      | Yes |    |

| Alcohol Free Test (g/210L) | 0.05g/210L Test (g/210L)<br>Lot#:201507A<br>Exp: 07/14/2017 | 0.08g/210L Test (g/210L)<br>Lot#:201601F<br>Exp: 01/26/2018 | 0.20g/210L Test (g/210L)<br>Lot#:201505A<br>Exp: 05/12/2017 | 0.08 g/210L Dry Gas Std Test (g/210L)<br>Lot#:03415080A1<br>Exp: 03/05/2017 |
|----------------------------|---|---|---|---|
| 0.000                      | 0.048   | 0.078   | 0.198   | 0.078   |
| 0.000                      | 0.049   | 0.078   | 0.200   | 0.078   |
| 0.000                      | 0.049   | 0.077   | 0.200   | 0.079   |

Number of Simulators Used: 5

Remarks:

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Ricky A. Frisco 3184 RICKY A FRISCO  
Signature and Printed Name

08/19/2016  
Date

# Florida Department of Law Enforcement Alcohol Testing Program

**RECEIVED**

JUL 07 2016

FDLE  
Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: PASCO COUNTY SO  
Time of Inspection: 09:54

Date of Inspection: 07/07/2016

Serial Number: 80-001081  
Software: 8100.27

| Check or Test                               | YES | NO |
|---|-----|----|
| Date and/or Time Adjusted                   |     | No |
| Diagnostic Check (Pre-Inspection): OK       | Yes |    |
| Alcohol Free Subject Test: 0.000            | Yes |    |
| Mouth Alcohol Test: Slope Not Met           | Yes |    |
| Interferent Detect Test: Interferent Detect | Yes |    |
| Diagnostic Check (Post-Inspection): OK      | Yes |    |

| Alcohol Free Test (g/210L) | 0.05g/210L Test (g/210L)<br>Lot#:201507A<br>Exp: 07/14/2017 | 0.08g/210L Test (g/210L)<br>Lot#:201601F<br>Exp: 01/26/2018 | 0.20g/210L Test (g/210L)<br>Lot#:201505A<br>Exp: 05/12/2017 | 0.08 g/210L Dry Gas Std Test (g/210L)<br>Lot#:03415080A1<br>Exp: 03/05/2017 |
|----------------------------|---|---|---|---|
| 0.000                      | 0.049   | 0.078   | 0.198   | 0.077   |
| 0.000                      | 0.049   | 0.078   | 0.199   | 0.077   |
| 0.000                      | 0.049   | 0.078   | 0.200   | 0.077   |

Number of Simulators Used: 5

Remarks:

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Ricky A Frisco 3184 RICKY A FRISCO  
Signature and Printed Name

07/07/2016  
Date

# Florida Department of Law Enforcement Alcohol Testing Program

**RECEIVED**

JUN 03 2016

FDLE  
Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: PASCO COUNTY SO  
Time of Inspection: 09:29

Date of Inspection: 06/03/2016

Serial Number: 80-001081  
Software: 8100.27

| Check or Test                               | YES | NO |
|---|-----|----|
| Date and/or Time Adjusted                   |     | No |
| Diagnostic Check (Pre-Inspection): OK       | Yes |    |
| Alcohol Free Subject Test: 0.000            | Yes |    |
| Mouth Alcohol Test: Slope Not Met           | Yes |    |
| Interferent Detect Test: Interferent Detect | Yes |    |
| Diagnostic Check (Post-Inspection): OK      | Yes |    |

| Alcohol Free Test (g/210L) | 0.05g/210L Test (g/210L)<br>Lot#:201507A<br>Exp: 07/14/2017 | 0.08g/210L Test (g/210L)<br>Lot#:201601F<br>Exp: 01/26/2018 | 0.20g/210L Test (g/210L)<br>Lot#:201505A<br>Exp: 05/12/2017 | 0.08 g/210L Dry Gas Std Test (g/210L)<br>Lot#:03415080A1<br>Exp: 03/05/2017 |
|----------------------------|---|---|---|---|
| 0.000                      | 0.050   | 0.077   | 0.198   | 0.079   |
| 0.000                      | 0.049   | 0.078   | 0.200   | 0.079   |
| 0.000                      | 0.050   | 0.078   | 0.201   | 0.079   |

Number of Simulators Used: 5

Remarks:

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

*Ricky A Frisco* 3184 RICKY A FRISCO  
Signature and Printed Name

06/03/2016  
Date

# Florida Department of Law Enforcement Alcohol Testing Program

RECEIVED  
MAY 05 2016  
FDLE  
Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: PASCO COUNTY SO  
Time of Inspection: 09:28

Date of Inspection: 05/05/2016

Serial Number: 80-001081  
Software: 8100.27

| Check or Test                               | YES | NO |
|---|-----|----|
| Date and/or Time Adjusted                   |     | No |
| Diagnostic Check (Pre-Inspection): OK       | Yes |    |
| Alcohol Free Subject Test: 0.000            | Yes |    |
| Mouth Alcohol Test: Slope Not Met           | Yes |    |
| Interferent Detect Test: Interferent Detect | Yes |    |
| Diagnostic Check (Post-Inspection): OK      | Yes |    |

| Alcohol Free Test (g/210L) | 0.05g/210L Test (g/210L)<br>Lot#:201507A<br>Exp: 07/14/2017 | 0.08g/210L Test (g/210L)<br>Lot#:201601F<br>Exp: 01/26/2018 | 0.20g/210L Test (g/210L)<br>Lot#:201505A<br>Exp: 05/12/2017 | 0.08 g/210L Dry Gas Std Test (g/210L)<br>Lot#:03415080A1<br>Exp: 03/05/2017 |
|----------------------------|---|---|---|---|
| 0.000                      | 0.048   | 0.077   | 0.200   | 0.079   |
| 0.000                      | 0.049   | 0.077   | 0.201   | 0.079   |
| 0.000                      | 0.049   | 0.078   | 0.201   | 0.078   |

Number of Simulators Used: 5

Remarks:

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Ricky A. Frisco 3184  
Signature and Printed Name

05/05/2016  
Date



# Florida Department of Law Enforcement Alcohol Testing Program

RECEIVED

APR 08 2016

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

FDLE  
Alcohol Testing Program

Agency: PASCO COUNTY SO  
Time of Inspection: 11:57

Date of Inspection: 04/08/2016

Serial Number: 80-001081  
Software: 8100.27

| Check or Test                               | YES | NO |
|---|-----|----|
| Date and/or Time Adjusted                   |     | No |
| Diagnostic Check (Pre-Inspection): OK       | Yes |    |
| Alcohol Free Subject Test: 0.000            | Yes |    |
| Mouth Alcohol Test: Slope Not Met           | Yes |    |
| Interferent Detect Test: Interferent Detect | Yes |    |
| Diagnostic Check (Post-Inspection): OK      | Yes |    |

| Alcohol Free Test (g/210L) | 0.05g/210L Test (g/210L)<br>Lot#:201507A<br>Exp: 07/14/2017 | 0.08g/210L Test (g/210L)<br>Lot#:201601F<br>Exp: 01/26/2018 | 0.20g/210L Test (g/210L)<br>Lot#:201505A<br>Exp: 05/12/2017 | 0.08 g/210L Dry Gas Std Test (g/210L)<br>Lot#:03415080A1<br>Exp: 03/05/2017 |
|----------------------------|---|---|---|---|
| 0.000                      | 0.049   | 0.078   | 0.199   | 0.080   |
| 0.000                      | 0.049   | 0.078   | 0.200   | 0.079   |
| 0.000                      | 0.050   | 0.078   | 0.200   | 0.079   |

Number of Simulators Used: 5

Remarks:

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Ricky A. Frisco 3184  
Signature and Printed Name

04/08/2016  
Date

RECEIVED

# Florida Department of Law Enforcement

## Alcohol Testing Program

 MAR 16 2016  
 FDLE  
 Alcohol Testing Program

### AGENCY INSPECTION REPORT - INTOXILYZER 8000

 Agency: PASCO COUNTY SO  
 Time of Inspection: 10:01

Date of Inspection: 03/16/2016

 Serial Number: 80-001081  
 Software: 8100.27

| Check or Test                               | YES | NO |
|---|-----|----|
| Date and/or Time Adjusted                   |     | No |
| Diagnostic Check (Pre-Inspection): OK       | Yes |    |
| Alcohol Free Subject Test: 0.000            | Yes |    |
| Mouth Alcohol Test: Slope Not Met           | Yes |    |
| Interferent Detect Test: Interferent Detect | Yes |    |
| Diagnostic Check (Post-Inspection): OK      | Yes |    |


| Alcohol Free Test (g/210L) | 0.05g/210L Test (g/210L)<br>Lot#:201507A<br>Exp: 07/14/2017 | 0.08g/210L Test (g/210L)<br>Lot#:201601F<br>Exp: 01/26/2018 | 0.20g/210L Test (g/210L)<br>Lot#:201505A<br>Exp: 05/12/2017 | 0.08 g/210L Dry Gas Std Test (g/210L)<br>Lot#:03415080A1<br>Exp: 03/05/2017 |
|----------------------------|---|---|---|---|
| 0.000                      | 0.050   | 0.078   | 0.200   | 0.079   |
| 0.000                      | 0.050   | 0.078   | 0.201   | 0.079   |
| 0.000                      | 0.052   | 0.078   | 0.201   | 0.079   |

Number of Simulators Used: 5

Remarks:

 The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

  
 3184 Signature and Printed Name RICKY A FRISCO

03/16/2016  
 Date

# Florida Department of Law Enforcement Alcohol Testing Program

**RECEIVED**  
FEB 29 2016  
FDLE  
Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: PASCO COUNTY SO  
Time of Inspection: 20:42

Date of Inspection: 02/25/2016

Serial Number: 80-001081  
Software: 8100.27

| Check or Test                               | YES | NO |
|---|-----|----|
| Date and/or Time Adjusted                   |     | No |
| Diagnostic Check (Pre-Inspection): OK       | Yes |    |
| Alcohol Free Subject Test: 0.000            | Yes |    |
| Mouth Alcohol Test: Slope Not Met           | Yes |    |
| Interferent Detect Test: Interferent Detect | Yes |    |
| Diagnostic Check (Post-Inspection): OK      | Yes |    |

| Alcohol Free Test (g/210L) | 0.05g/210L Test (g/210L)<br>Lot#:201507A<br>Exp: 07/14/2017 | 0.08g/210L Test (g/210L)<br>Lot#:201502G<br>Exp: 02/24/2017 | 0.20g/210L Test (g/210L)<br>Lot#:201505A<br>Exp: 05/12/2017 | 0.08 g/210L Dry Gas Std Test (g/210L)<br>Lot#:0341508A1<br>Exp: 03/05/2017 |
|----------------------------|---|---|---|--|
| 0.000                      | 0.050   | 0.078   | 0.200   | 0.079  |
| 0.000                      | 0.050   | 0.078   | 0.201   | 0.080  |
| 0.000                      | 0.051   | 0.078   | 0.201   | 0.079  |

Number of Simulators Used: 5

**Remarks:**

A F / M A: PROVIDED 2 ALCH FREE SAMPLES.

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Ricky A Frisco 3184 RICKY A FRISCO  
Signature and Printed Name

02/25/2016  
Date