

Florida Department of Law Enforcement SEP 16 2016
 Alcohol Testing Program FDLE
Alcohol Testing Program
 AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: EGLIN AFB
 Time of Inspection: 13:04

Date of Inspection: 08/17/2016

Serial Number: 80-001116
 Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: _____ Exp: _____	0.08g/210L Test (g/210L) Lot#: _____ Exp: _____	0.20g/210L Test (g/210L) Lot#: _____ Exp: _____	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: _____ Exp: _____

Number of Simulators Used: _____

Remarks:
 SKIPPED AI TO OPERATE INSTRUMENT; NOT A COMPLIANCE CHECK

DM
BK

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Phillip W Suber

PHILLIP W SUBER

Signature and Printed Name

08/17/2016
 Date

Florida Department of Law Enforcement Alcohol Testing Program

RECEIVED

SEP 16 2016

FDLE

AGENCY INSPECTION REPORT - INTOXILYZER 8000 Alcohol Testing Program

Agency: EGLIN AFB
Time of Inspection: 10:31

Date of Inspection: 09/14/2016

Serial Number: 80-001116
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: _____ Exp: _____	0.08g/210L Test (g/210L) Lot#: _____ Exp: _____	0.20g/210L Test (g/210L) Lot#: _____ Exp: _____	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: _____ Exp: _____

PJM

Number of Simulators Used: _____

Remarks:
SKIPPED AI TO OPERATE INSTRUMENT; NOT A COMPLIANCE CHECK

BK

The above instrument complies (*N/A*) does not comply (*(M/S)*) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

[Handwritten Signature]

PHILLIP W SUBER

Signature and Printed Name

09/14/2016
Date