

Florida Department of Law Enforcement Alcohol Testing Program

RECEIVED

JUN 01 2016

FDLE
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: FL HIGHWAY PATROL
Time of Inspection: 11:31

Date of Inspection: 06/01/2016

Serial Number: 80-001118
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	


Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201507A Exp: 07/14/2017	0.08g/210L Test (g/210L) Lot#:201502G Exp: 02/24/2017	0.20g/210L Test (g/210L) Lot#:201505A Exp: 05/12/2017	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:28814080A3 Exp: 11/05/2016
0.000	0.048	0.078	0.196	0.079
0.000	0.048	0.078	0.197	0.078
0.000	0.048	0.078	0.197	0.079

Number of Simulators Used: 5

Remarks:
NA

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



SUSAN M BARGE

Signature and Printed Name

06/01/2016
Date

RECEIVED

MAY 03 2016

FDLE Alcohol Testing Program

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: FL HIGHWAY PATROL
Time of Inspection: 15:44

Date of Inspection: 05/03/2016

Serial Number: 80-001118
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	


Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201507A Exp: 07/14/2017	0.08g/210L Test (g/210L) Lot#:201502G Exp: 02/24/2017	0.20g/210L Test (g/210L) Lot#:201505A Exp: 05/12/2017	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:28814080A3 Exp: 11/05/2015
0.000	0.048	0.077	0.199	0.079
0.000	0.048	0.077	0.198	0.079
0.000	0.049	0.077	0.199	0.079

Number of Simulators Used: 5

Remarks:
NONE

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



 Signature and Printed Name SUSAN M BARCE

05/03/2016
Date

RECEIVED

Florida Department of Law Enforcement

Alcohol Testing Program

 APR 06 2016
 Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

 Agency: FL HIGHWAY PATROL
 Time of Inspection: 10:13

Date of Inspection: 04/06/2016

 Serial Number: 80-001118
 Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	


Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201507A Exp: 07/14/2017	0.08g/210L Test (g/210L) Lot#: 201502G Exp: 02/24/2017	0.20g/210L Test (g/210L) Lot#: 201505A Exp: 05/12/2017	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 07014080A2 Exp: 05/01/2016
0.000	0.049	0.078	0.199	0.080
0.000	0.049	0.078	0.199	0.080
0.000	0.049	0.078	0.199	0.080

 Number of Simulators Used: 5

 Remarks:
 N/A

 The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



SUSAN M. BARGE

Signature and Printed Name

04/06/2016
 Date

80-001118

Florida Department of Law Enforcement Alcohol Testing Program

RECEIVED
FEB 09 2016

AGENCY INSPECTION REPORT - INTOXILYZER 8000

FDLE

Agency: FL HIGHWAY PATROL
Time of Inspection: 15:15

Date of Inspection: 02/01/2016

Serial Number: 80-001118
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201502A Exp: 02/19/2016	0.08g/210L Test (g/210L) Lot#: 201502B Exp: 02/24/2017	0.20g/210L Test (g/210L) Lot#: 201502A Exp: 05/12/2017	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 201502A Exp: 02/19/2016
0.000	0.049	0.078	0.198	0.079
0.000	0.049	0.079	0.198	0.079
0.000	0.048	0.079	0.198	0.080

Number of Simulators Used: 3

Remarks:
NONE

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



Signature and Printed Name: GREGORY J. LAMBERT

02/01/2016
Date

RECEIVED

Florida Department of Law Enforcement JAN 29 2016
 Alcohol Testing Program

FDLE
 Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: FL HIGHWAY PATROL
 Time of Inspection: 16:46

Date of Inspection: 01/28/2016

Serial Number: 80-001118
 Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	


Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 261507A Exp: 02/14/2016	0.08g/210L Test (g/210L) Lot#: 261502G Exp: 02/24/2017	0.20g/210L Test (g/210L) Lot#: 261505A Exp: 05/12/2017	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 07014020A3 Exp: 05/01/2016
0.000	0.046	0.077	0.197	0.079
0.000	0.048	0.077	0.197	0.078
0.000	0.048	0.077	0.197	0.079

Number of Simulators Used: 5

Remarks:
 NONE

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



Signature and Printed Name: GREGORY E. LANONI

Date: 01/28/2016