

RECEIVED

MAY 08 2016

FDLE  
Alcohol Testing Program

# Florida Department of Law Enforcement Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: FL HIGHWAY PATROL  
Time of Inspection: 16:02

Date of Inspection: 05/03/2016

Serial Number: 80-001119  
Software: 8100.27

| Check or Test                               | YES | NO |
|---|-----|----|
| Date and/or Time Adjusted                   |     | No |
| Diagnostic Check (Pre-Inspection): OK       | Yes |    |
| Alcohol Free Subject Test: 0.000            | Yes |    |
| Mouth Alcohol Test: Slope Not Met           | Yes |    |
| Interferent Detect Test: Interferent Detect | Yes |    |
| Diagnostic Check (Post-Inspection): OK      | Yes |    |

| Alcohol Free Test (g/210L) | 0.05g/210L Test (g/210L)<br>Lot#: 201507A<br>Exp: 07/14/2017 | 0.08g/210L Test (g/210L)<br>Lot#: 201502G<br>Exp: 02/24/2017 | 0.20g/210L Test (g/210L)<br>Lot#: 201505A<br>Exp: 05/12/2017 | 0.08 g/210L Dry Gas Std Test (g/210L)<br>Lot#: 26814080A3<br>Exp: 11/05/2016 |
|----------------------------|--|--|--|--|
| 0.000                      | 0.050  | 0.078  | 0.201  | 0.078  |
| 0.000                      | 0.050  | 0.079  | 0.201  | 0.078  |
| 0.000                      | 0.050  | 0.079  | 0.201  | 0.078  |

Number of Simulators Used: 5


Remarks:

A F / M A: NOT DEEP LUNG AIR



The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



SUSAN M BARGE

Signature and Printed Name.

05/03/2016  
Date