

RECEIVED

DEC 07 2016

Florida Department of Law Enforcement Alcohol Testing Program

FDLE
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: Broward County S.O.

Serial Number: 80-001139

Time of Inspection:

Date of Inspection: 11/20/2016

Software:

CHECK OR TEST	YES	NO
Date and/or Time Adjusted		
Diagnostic Check (Pre-Inspection): <u>OK</u>		
Alcohol Free Subject Test: <u>0.000</u>		
Mouth Alcohol Test: <u>Slope Not Met</u>		
Interferent Detect Test: <u>Interferent Detect</u>		
Diagnostic Check (Post-Inspection): <u>OK</u>		

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: _____ Exp: _____	0.08g/210L Test (g/210L) Lot#: _____ Exp: _____	0.20g/210L Test (g/210L) Lot#: _____ Exp: _____	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: _____ Exp: _____

Number of Simulators Used:

Remarks:

Unable to do inspection due to instrument needs to be shipped for regulator replacement. Approval for shipping not in budget yet. Will be shipped as soon as approval for repairs and shipping is received.

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Laura R. D. Cery

 Signature and Printed Name

11/20/2016

 Date

RECEIVED

DEC 07 2016

Florida Department of Law Enforcement Alcohol Testing Program

FDLE
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: Broward County S.O.

Time of Inspection:

Date of Inspection: 10/19/2016

Serial Number: 80-601139

Software:

CHECK OR TEST	YES	NO
Date and/or Time Adjusted		
Diagnostic Check (Pre-Inspection): OK		
Alcohol Free Subject Test: 0.000		
Mouth Alcohol Test: Slope Not Met		
Interferent Detect Test: Interferent Detect		
Diagnostic Check (Post-Inspection): OK		

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: _____ Exp: _____	0.08g/210L Test (g/210L) Lot#: _____ Exp: _____	0.20g/210L Test (g/210L) Lot#: _____ Exp: _____	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: _____ Exp: _____

Number of Simulators Used: 5

Remarks:

Unable to perform Inspection due to instrument still packaged to ship for new regulator installment. Needs to wait budget approval to ship.

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Francis R. DiCenzo 10/19/2016
Signature and Printed Name

10/19/2016
Date

RECEIVED

MAR 02 2016

Florida Department of Law Enforcement Alcohol Testing Program

FDLE
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: BROWARD COUNTY SO
Time of Inspection: 21:11

Date of Inspection: 01/29/2016

Serial Number: 80-001139
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201507A Exp: 07/14/2017	0.08g/210L Test (g/210L) Lot#:201502G Exp: 02/24/2017	0.20g/210L Test (g/210L) Lot#:201505A Exp: 05/12/2017	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:1789170 Exp: 01/14/2018
0.000	0.051	0.079	0.201	0.079
0.000	0.051	0.080	0.203	0.079
0.000	0.052	0.079	0.204	0.079

Number of Simulators Used: 5

Remarks:

T. SUN WAS PRESENT AND OBSERVED ENTIRE INSPECTION.

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Franca R. Dicienzo

FRANCA R DICIENZO

Signature and Printed Name

01/29/2016
Date