

**AMENDED**  
 Florida Department of Law Enforcement  
 Alcohol Testing Program

**RECEIVED**  
 JUL 28 2016  
 FDLE  
 Alcohol Testing Program

**AGENCY INSPECTION REPORT - INTOXILYZER 8000**

Agency: HAINES CITY PD Serial Number: 80-001151  
 Time of Inspection: 06:20 Date of Inspection: 06/27/2016 Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: _____ Exp: _____	0.08g/210L Test (g/210L) Lot#: _____ Exp: _____	0.20g/210L Test (g/210L) Lot#: _____ Exp: _____	0.08 g/210L Dry Gas Std Test (g/210L) Lot #: _____ Exp: _____

Number of Simulators Used: \_\_\_\_\_

Remarks: WHEN I WENT TO CONDUCT THE INSPECTION, I ACCIDENTALLY PRESSED THE #2 KEY TWICE INSTEAD OF THE ESC KEY WHICH IN TURN MESS'D EVERYTHING UP. I ATTEMPTED TO ABORT THE INSPECTION & WAS UNABLE TO. I THEN QUICKLY PRESSED THE RETURN KEY REPEATABLY IN ORDER TO QUICKLY GO THROUGH THE PROMPTS SO I CAN RESTART THE INSPECTION.

The above instrument complies ( ~~X~~ ) does not comply (  ) with Chapter 11D-8, FAC.  
*Amended 07/28/16*

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



RAY M WEBER

Signature and Printed Name

06/27/2016

Date