

# Florida Department of Law Enforcement Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: MIAMI PD  
Time of Inspection: 13:01

Date of Inspection: 08/11/2016

Serial Number: 80-001178  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 0111070 Exp: 07-11-2017	0.08g/210L Test (g/210L) Lot#: 0111070 Exp: 07-11-2017	0.20g/210L Test (g/210L) Lot#: 0111070 Exp: 07-11-2017	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 1111070 Exp: 07-11-2017
0.000	0.052	0.052	0.203	0.050
0.000	0.052	0.052	0.203	0.050
0.000	0.052	0.052	0.202	0.050

Number of Simulators Used: 0

Remarks: ISA B Reyes Present

**RECEIVED**

SEP 23 2016

**FDLE  
Alcohol Testing Program**

The above instrument complies ( X ) does not comply ( ) with Chapter 11D-8, F.A.C.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, F.A.C.

28394

Signature and Printed Name

Date