

RECEIVED

JAN 05 2016

Florida Department of Law Enforcement Alcohol Testing Program

FDLE
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: POLK COUNTY SO
Time of Inspection: 12:25

Date of Inspection: 01/05/2016

Serial Number: 80-001237
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201507A Exp: 07/14/2017	0.08g/210L Test (g/210L) <i>2G BW</i> Lot#:201507A Exp: 07/14/2017 <i>2124</i>	0.20g/210L Test (g/210L) Lot#:201505A Exp: 05/12/2017	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:07014080A2 Exp: 05/01/2016
0.000	0.048	0.079	0.202	0.079
0.000	0.048	0.079	0.201	0.079
0.000	0.048	0.079	0.200	0.079

Number of Simulators Used: 5

Remarks:

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

6493
Signature and Printed Name

BARBIE R WILSON

01/05/2016
Date