AMEND

Florida Department of Law Enforcement
Alcohol Testing Program

1242

RECEIVED
JUN 17 2016
FDLE
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: DAYTONA BEACH SHORES
Time of Inspection: 05:26 Date of Inspection: 06/07/2016

Serial Number: 80-001242
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201507A Exp: 07/14/2017	0.08g/210L Test (g/210L) Lot#:201502G Exp: 02/24/2017	0.20g/210L Test (g/210L) Lot#:201505A Exp: 05/12/2017	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: Exp:
0.000	0.048	0.077	0.170 / 0.189	
0.000	0.048	0.077	0.191 / 0.197	
0.000	0.049	0.078	0.193 / 0.197	

Number of Simulators Used: 4

Remarks:
20: Control Outside Tolerance, Control Outside Tolerance.

CALLED BY SGT. ON RADIO WHILE CONNECTING 0.08 GAS.
MUST HAVE HIT NO BY ACCIDENT WHEN ASKED.

6-7-16 @ 0527 DGM

* NOT IN COMPLIANCE *

The above instrument complies (X) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

David G Murphey — DAVID G MURPHEY SR
Signature and Printed Name

06/07/2016
Date

FDLE/ATP Form 40 – March 2004