

Florida Department of Law Enforcement Alcohol Testing Program

RECEIVED

JUL 05 2016

FDLE

AGENCY INSPECTION REPORT - INTOXILYZER 8000 Alcohol Testing Program

Agency: BAKER COUNTY SO
Time of Inspection: 08:20

Date of Inspection: 06/29/2016

Serial Number: 80-001287
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201507A Exp: 07/14/2017	0.08g/210L Test (g/210L) Lot#:201502G Exp: 02/24/2017	0.20g/210L Test (g/210L) Lot#:201505A Exp: 05/12/2017	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:24515080A1 Exp: 10/05/2017
0.000	0.048	0.077	0.201	0.081
0.000	0.048	0.078	0.202	0.081
0.000	0.048	0.078	0.199	0.081


Number of Simulators Used: 5

Remarks:

**BAKER COUNTY
SHERIFF'S OFFICE
JUN 29 2016
RECORDS SECTION
RECEIVED**

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

 COREY S WILSON

Signature and Printed Name

06/29/2016
Date