

RECEIVED

Florida Department of Law Enforcement
Alcohol Testing Program

DEC 07 2016

FDLE
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: BROWARD COUNTY SO
Time of Inspection: 19:08

Date of Inspection: 11/20/2016

Serial Number: 80-001248
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201507A Exp: 07/14/2017	0.08g/210L Test (g/210L) Lot#:201502G Exp: 02/24/2017	0.20g/210L Test (g/210L) Lot#:201505A Exp: 05/12/2017	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:34015080A3 Exp: 01/05/2018
0.000	0.050	0.078	0.199	0.081
0.000	0.049	0.078	0.200	0.081
0.000	0.049	0.079	0.202	0.081

Number of Simulators Used: 5

Remarks:

A. FRAZIER & Y. WASHINGTON OBSERVED ENTIRE INSPECTION

The above instrument complies (X) does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Franca R. Dicienzo FRANCA R DICIENZO
Signature and Printed Name

11/20/2016
Date

RECEIVED

Florida Department of Law Enforcement
 Alcohol Testing Program

SEP 27 2016

FDLE
 Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: BROWARD COUNTY SO
 Time of Inspection: 21:00

Date of Inspection: 09/24/2016

Serial Number: 80-001248
 Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201507A Exp: 07/14/2017	0.08g/210L Test (g/210L) Lot#:201502G Exp: 02/24/2017	0.20g/210L Test (g/210L) Lot#:201505A Exp: 05/12/2017	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:34015080A3 Exp: 01/05/2018
0.000	0.050	0.079	0.201	0.080
0.000	0.050	0.079	0.202	0.080
0.000	0.051	0.079	0.202	0.080

Number of Simulators Used: 5

Remarks:

A. FRAZIER OBSERVED & Y WASHINGTON IN TRAINING

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Franca R. Dicienzo FRANCA R DICIENZO
 Signature and Printed Name

09/24/2016
 Date

RECEIVED

MAR 02 2016

Florida Department of Law Enforcement Alcohol Testing Program

FDLE
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: BROWARD COUNTY SO
Time of Inspection: 21:12

Date of Inspection: 01/29/2016

Serial Number: 80-001248
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201507A Exp: 07/14/2017	0.08g/210L Test (g/210L) Lot#: 201502G Exp: 02/24/2017	0.20g/210L Test (g/210L) Lot#: 201505A Exp: 05/12/2017	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 15615080A3 Exp: 07/05/2017
0.000	0.052	0.081	0.205	0.081
0.000	0.051	0.080	0.206	0.080
0.000	0.051	0.080	0.205	0.080


Number of Simulators Used: 5

Remarks:

F DICIZENZO WAS PRESENT & OBSERVED ENTIRE INSPECTION

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.


Signature and Printed Name

TONY SUN

01/29/2016
Date