

RECEIVED

DEC 07 2016

Florida Department of Law Enforcement Alcohol Testing Program

FDLE
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: BROWARD COUNTY SO
Time of Inspection: 20:47

Date of Inspection: 11/20/2016

Serial Number: 80-001249
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201507A Exp: 07/14/2017	0.08g/210L Test (g/210L) Lot#: 201502G Exp: 02/24/2017	0.20g/210L Test (g/210L) Lot#: 201505A Exp: 05/12/2017	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 34015080A3 Exp: 01/05/2018
0.000	0.049	0.078	0.199	0.082
0.000	0.050	0.079	0.201	0.082
0.000	0.049	0.079	0.200	0.081

Number of Simulators Used: 5

Remarks:

A PRAZIER & Y WASHINGTON OBS 2ND INSP. O RINGS & GAS CHANGE

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

FRANCA R DICIENZO

Signature and Printed Name

11/20/2016

Date

RECEIVED

DEC 07 2016

Florida Department of Law Enforcement Alcohol Testing Program

FDLE
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: BROWARD COUNTY SO
Time of Inspection: 19:54

Date of Inspection: 11/20/2016

Serial Number: 80-001249
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201507A Exp: 07/14/2017	0.08g/210L Test (g/210L) Lot#: 201507A Exp: 07/14/2017 <i>201507A</i> <i>02/04/2017</i>	0.20g/210L Test (g/210L) Lot#: 201505A Exp: 05/12/2017	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 34015080A3 Exp: 01/05/2018
0.000	0.049	0.078	0.203	0.090 / 0.086
0.000	0.049	0.079	0.201	0.087 / 0.085
0.000	0.050	0.079	0.201	0.086 / 0.085

Number of Simulators Used: 5

Remarks:

08: Control Outside Tolerance, FRAZIER & WASHINGTON OBS . OUTSIDE TOLERANCE, UNK FRAZIER & WAHSHIN GTON OBSERVED Non-compliance: .08 GAS OUTSIDE TOLERANCE.

The above instrument complies () does not comply (X) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Franca R. Dicienzo

Signature and Printed Name

11/20/2016
Date

RECEIVED

SEP 27 2016

Florida Department of Law Enforcement Alcohol Testing Program

FDLE
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: BROWARD COUNTY SO
Time of Inspection: 21:33

Date of Inspection: 09/24/2016

Serial Number: 80-001249
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201507A Exp: 07/14/2017	0.08g/210L Test (g/210L) Lot#:201502G Exp: 02/24/2017	0.20g/210L Test (g/210L) Lot#:201505A Exp: 05/12/2017	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:34015080A3 Exp: 01/05/2018
0.000	0.049	0.079	0.199	0.076
0.000	0.049	0.080	0.201	0.076
0.000	0.049	0.080	0.200	0.076

Number of Simulators Used: 5

Remarks:

A. FRAZIER OBSERVED & Y WASHINGTON IN TRAINING

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



FRANCA R DICIENZO

Signature and Printed Name

09/24/2016
Date

Florida Department of Law Enforcement MAR 15 2016
 Alcohol Testing Program FDLE
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: BROWARD COUNTY SO
 Time of Inspection: 09:20

Date of Inspection: 03/07/2016

Serial Number: 80-001249
 Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201507A Exp: 07/14/2017	0.08g/210L Test (g/210L) Lot#:201502G Exp: 02/24/2017	0.20g/210L Test (g/210L) Lot#:201505A Exp: 05/12/2017	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:1789170 Exp: 01/14/2018
0.000	0.051	0.081	0.202	0.081
0.000	0.051	0.082	0.202	0.081
0.000	0.052	0.081	0.202	0.081

Number of Simulators Used: 5

Remarks:
 LAB TESTING OF INSTRUMENT

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

David E Reyes-Rivera

DAVID E REYES-RIVERA

Signature and Printed Name

03/07/2016
 Date

RECEIVED

MAR 02 2016

Florida Department of Law Enforcement Alcohol Testing Program

FDLE
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: BROWARD COUNTY SO
Time of Inspection: 19:43

Date of Inspection: 01/29/2016

Serial Number: 80-001249
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201507A Exp: 07/14/2017	0.08g/210L Test (g/210L) Lot#:201502G Exp: 02/24/2017	0.20g/210L Test (g/210L) Lot#:201505A Exp: 05/12/2017	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:1789170 Exp: 01/14/2018
0.000	0.050	0.079	0.201	0.081
0.000	0.051	0.079	0.203	0.081
0.000	0.051	0.079	0.203	0.081

Number of Simulators Used: 5

Remarks:

T.SUN WAS PRESENT AND OBSERVED ENTIRE INSPECTION.

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Franca R. Dicienzo

FRANCA R DICIZENZO

Signature and Printed Name

01/29/2016
Date