

# Florida Department of Law Enforcement Alcohol Testing Program

**RECEIVED**

FEB 26 2016 ACC

*Open*

Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: BROWARD COUNTY SO  
Time of Inspection: 08:27

Date of Inspection: 02/09/2016

Serial Number: 80-001250  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted	Yes	
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: _____ Exp: _____	0.08g/210L Test (g/210L) Lot#: _____ Exp: _____	0.20g/210L Test (g/210L) Lot#: _____ Exp: _____	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: _____ Exp: _____

Number of Simulators Used: \_\_\_\_\_

**Remarks:**

Time-Date changed. ADJUSTED

*Compliance not determined. AI not conducted, by passed to get instrument out of Disabled mode.*

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

*David E Reyes Rivera*

DAVID E REYES-RIVERA

Signature and Printed Name

02/09/2016  
Date