

# Florida Department of Law Enforcement Alcohol Testing Program

**RECEIVED**

DEC 14 2016

FDLE

## AGENCY INSPECTION REPORT - INTOXILYZER 8000 Alcohol Testing Program

Agency: ROCKLEDGE P.D.

Time of Inspection: 00:28

Date of Inspection: 11/30/2016

Serial Number: 80-001261

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted	Yes	
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201507A Exp: 07/14/2017	0.08g/210L Test (g/210L) Lot#:201601F Exp: 01/26/2018	0.20g/210L Test (g/210L) Lot#:201505A Exp: 05/12/2017	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:572920 Exp: 09/13/2019
0.000	0.050	0.079	0.200	0.079
0.000	0.050	0.080	0.201	0.080
0.000	0.050	0.080	0.201	0.080

Number of Simulators Used: 4

**Remarks:**

Time-Date changed. A F / M A: Improper Sample.

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

*Patrick W Kennedy* PATRICK W KENNEDY  
Signature and Printed Name

11/30/2016  
Date

RECEIVED

# Florida Department of Law Enforcement Alcohol Testing Program

SEP 06 2016

FDLE

Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: ROCKLEDGE P.D.  
Time of Inspection: 23:32

Date of Inspection: 08/29/2016

Serial Number: 80-001261  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201507A Exp: 07/14/2017	0.08g/210L Test (g/210L) Lot#:201601F Exp: 01/26/2018	0.20g/210L Test (g/210L) Lot#:201505A Exp: 05/12/2017	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:28214080A1 Exp: 11/05/2016
0.000	0.050	0.079	0.201	0.078
0.000	0.050	0.080	0.202	0.078
0.000	0.050	0.081	0.202	0.078

Number of Simulators Used: 4

Remarks:

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

*Sgt. Ivette Gomez #130*

IVETTE GOMEZ

Signature and Printed Name

08/29/2016  
Date

AUG 04 2016

# Florida Department of Law Enforcement Alcohol Testing Program

FDLE  
Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: ROCKLEDGE P.D.  
Time of Inspection: 21:28

Date of Inspection: 07/28/2016

Serial Number: 80-001261  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	


Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201507A Exp: 07/14/2017	0.08g/210L Test (g/210L) Lot#:201601F Exp: 01/26/2018	0.20g/210L Test (g/210L) Lot#:201505A Exp: 05/12/2017	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:28214080A1 Exp: 11/05/2016
0.000	0.050	0.079	0.199	0.079
0.000	0.050	0.080	0.200	0.079
0.000	0.050	0.080	0.201	0.079

Number of Simulators Used: 4

Remarks:

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.


 #180

IVETTE GOMEZ

Signature and Printed Name

07/28/2016  
Date

# Florida Department of Law Enforcement Alcohol Testing Program

  
**RECEIVED**  
 JUL 20 2016  
 JUL 06 2016  
 FILE  
 Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: ROCKLEDGE P.D.  
Time of Inspection: 12:43

Date of Inspection: 06/28/2016

Serial Number: 80-001261  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	


Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201507A Exp: 07/14/2017	0.08g/210L Test (g/210L) Lot#:201601F Exp: 01/26/2018	0.20g/210L Test (g/210L) Lot#:201505A Exp: 05/12/2017	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:28214080A1 Exp: 11/05/2016
0.000	0.050	0.080	0.201	0.079
0.000	0.050	0.080	0.200	0.079
0.000	0.050	0.080	0.201	0.079

Number of Simulators Used: 4

Remarks:

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

  
 Signature and Printed Name PATRICK W KENNEDY

06/28/2016  
Date

RECEIVED

Florida Department of Law Enforcement  
 Alcohol Testing Program

MAY 26 2016

FDLE

Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: ROCKLEDGE P.D.  
 Time of Inspection: 04:30

Date of Inspection: 05/18/2016

Serial Number: 80-001261  
 Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201507A Exp: 07/14/2017	0.08g/210L Test (g/210L) Lot#:201502G Exp: 02/24/2017	0.20g/210L Test (g/210L) Lot#:201505A Exp: 05/12/2017	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:28214080A1 Exp: 11/05/2016
0.000	0.050	0.080	0.202	0.078
0.000	0.050	0.080	0.202	0.078
0.000	0.050	0.080	0.202	0.078

Number of Simulators Used: 4

Remarks:

The above instrument complies ( X ) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

*SGT. Patrick W. Kennedy #185*  
 \_\_\_\_\_  
 Signature and Printed Name

PATRICK W KENNEDY

05/18/2016  
 Date

RECEIVED

# Florida Department of Law Enforcement Alcohol Testing Program

MAY 05 2016

FDLE  
Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: ROCKLEDGE P.D.  
Time of Inspection: 05:46

Date of Inspection: 04/29/2016

Serial Number: 80-001261  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201507A Exp: 07/14/2017	0.08g/210L Test (g/210L) Lot#:201502G Exp: 02/24/2017	0.20g/210L Test (g/210L) Lot#:201505A Exp: 05/12/2017	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:28214080A1 Exp: 11/05/2016
0.000	0.050	0.080	0.201	0.078
0.000	0.050	0.080	0.201	0.078
0.000	0.050	0.080	0.202	0.078

Number of Simulators Used: 4

Remarks:

COPY

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

567 Patrick W Kennedy #185 PATRICK W KENNEDY  
Signature and Printed Name

04/29/2016  
Date

1261

# Florida Department of Law Enforcement Alcohol Testing Program

**RECEIVED**

FEB 23 2016

FDLE

## AGENCY INSPECTION REPORT - INTOXILYZER 8000 Alcohol Testing Program

Agency: ROCKLEDGE P.D.  
Time of Inspection: 00:29

Date of Inspection: 02/11/2016

Serial Number: 80-001261  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201507A Exp: 07/14/2017	0.08g/210L Test (g/210L) Lot#:201502G Exp: 02/24/2017	0.20g/210L Test (g/210L) Lot#:201505A Exp: 05/12/2017	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:28214080A1 Exp: 11/05/2016
0.000	0.052	0.081	0.202	0.079
0.000	0.052	0.081	0.203	0.079
0.000	0.052	0.082	0.204	0.079

Number of Simulators Used: 4

Remarks:

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

*Sgt. Patrick W. Kennedy #785*  
 \_\_\_\_\_  
 Signature and Printed Name

02/11/2016  
Date

# Florida Department of Law Enforcement Alcohol Testing Program

1261  
RECEIVED  
JAN 20 2016  
FDLE  
Alcohol Testing Program  
COPY

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: ROCKLEDGE P.D.  
Time of Inspection: 04:59

Date of Inspection: 01/14/2016

Serial Number: 80-001261  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201507A Exp: 07/14/2017	0.08g/210L Test (g/210L) Lot#:201502G Exp: 02/24/2017	0.20g/210L Test (g/210L) Lot#:201505A Exp: 05/12/2017	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:2814080A1 Exp: 11/05/2016
0.000	0.051	0.080	0.202	0.079
0.000	0.051	0.081	0.203	0.079
0.000	0.051	0.081	0.204	0.079

Number of Simulators Used: 4

Remarks:

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

  
 \_\_\_\_\_ PATRICK W KENNEDY  
 Signature and Printed Name

01/14/2016  
Date