

RECEIVED

# Florida Department of Law Enforcement Alcohol Testing Program

MAY 24 2016

FDLE

Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: BREVARD COUNTY S.O.      Date of Inspection: 05/23/2016      Serial Number: 80-001264  
 Time of Inspection: 09:03      Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: _____ Exp: _____	0.08g/210L Test (g/210L) Lot#: _____ Exp: _____	0.20g/210L Test (g/210L) Lot#: _____ Exp: _____	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: _____ Exp: _____


*JLM*

Number of Simulators Used: \_\_\_\_\_

Remarks:  
BYPASS AI TO OPERATE INSTRUMENTYNon-compliance:NOT DETERMINED.

The above instrument complies ( ) does not comply ( X ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



JAKE L SHANAHAN

Signature and Printed Name

05/23/2016  
Date

# Florida Department of Law Enforcement Alcohol Testing Program

RECEIVED

MAY 24 2016

FDLE

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Alcohol Testing Program

Agency: BREVARD COUNTY S.O.  
Time of Inspection: 09:03

Date of Inspection: 05/23/2016

Serial Number: 80-001264  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: _____ Exp: _____	0.08g/210L Test (g/210L) Lot#: _____ Exp: _____	0.20g/210L Test (g/210L) Lot#: _____ Exp: _____	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: _____ Exp: _____


*gjm*

Number of Simulators Used: \_\_\_\_\_

Remarks:  
BYPASS AI TO OPERATE INSTRUMENT Non-compliance: NOT DETERMINED.

The above instrument complies (     ) does not comply ( X ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



JAKE L SHANAHAN

Signature and Printed Name

05/23/2016  
Date