

RECEIVED

SEP 28 2016

1265

FDLE
Alcohol Testing Program

Florida Department of Law Enforcement
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: PALM BAY P.D.

Serial Number: 80-001265

Time of Inspection: 11:33

Date of Inspection: 09/25/2016

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201507A Exp: 07/14/2017	0.08g/210L Test (g/210L) Lot#:201601F Exp: 01/26/2018	0.20g/210L Test (g/210L) Lot#:201505A Exp: 05/12/2017	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:24515080A2 Exp: 10/05/2017
0.000	0.049	0.080	0.201	0.080
0.000	0.049	0.082	0.201	0.079
0.000	0.050	0.081	0.201	0.080

Number of Simulators Used: 5

Remarks:

The above instrument complies (X) does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

EDWIN J LUTZ
Signature and Printed Name

09/25/2016
Date

1265
P

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: PALM BAY P.D.
Time of Inspection: 10:29

Date of Inspection: 08/23/2016

Serial Number: 80-001265
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	


Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201507A Exp: 07/14/2017	0.08g/210L Test (g/210L) Lot#:201601F Exp: 01/26/2018	0.20g/210L Test (g/210L) Lot#:201505A Exp: 05/12/2017	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:24515080A2 Exp: 10/05/2017
0.000	0.051	0.080	0.201	0.080
0.000	0.050	0.081	0.202	0.079
0.000	0.050	0.081	0.202	0.080

Number of Simulators Used: 5

Remarks:

The above instrument complies () does not comply () with Chapter 11D-8, FAC.


I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



EDWIN J LUTZ

Signature and Printed Name

08/23/2016
Date

 D. Morgan #46

Florida Department of Law Enforcement Alcohol Testing Program

RECEIVED

JUN 29 2016

FDLE

Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: PALM BAY P.D.
Time of Inspection: 16:14

Date of Inspection: 06/27/2016

Serial Number: 80-001265
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201507A Exp: 07/14/2017	0.08g/210L Test (g/210L) Lot#: Exp:	0.20g/210L Test (g/210L) Lot#: Exp:	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: Exp:
0.000	0.050 / INT			
0.000	INT			
0.000				

Number of Simulators Used: 5

Remarks:


INT DETECT / UNKNOWN ERROR. UNKNOWN INT DETECT ; PURGE FAIL OCCURRED.

INSTRUMENT EXPERIENCED UNKNOWN ERROR DURING .05 SIM.
2ND SAMPLE RETURNED INT DETECT; SUBSEQUENT PURGE FAIL. REPEATED, INT DETECT STILL PRESENT.

TAKEN OUT OF SERVICE; DEPT INSPECTOR J. SHAWANAN NOTIFIED. BEING RETURNED TO A.T.P.

The above instrument complies () does not comply (X) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



EDWIN J LUTZ

Signature and Printed Name

06/27/2016
Date

JUN 30 2016

FDLE
Alcohol Testing Program

JAKE, THIS ONE IS
IN A PATROL CAR (MINE). I
NOTICED A COUPLE DAYS AGO IT
GAVE ME AN INT DETECT, BUT
I COULD SMELL EXHAUST FUMES AND
SUSPECTED IT WAS FROM THE VEHICLE'S
EXHAUST.

DURING THE INSPECTION TODAY,
GAVE ME THE INT. DETECT AND
THEN A PURGE FAIL. WHEN I REPEATED,
THE INSTRUMENT WAS REGISTERING THE
.05 SOLUTION AT A .015 - .016 AND
THEN GAVE ME ANOTHER INT. DETECT,
HOWEVER NO PURGE FAIL THIS TIME.

I'D SUSPECT SOME INT SOLUTION GOT
INTO THE INSTRUMENT OR THERE IS
SOME INTERNAL FAILURE.

- E LUTZ, PALM BAY P.D.

1265
Ⓟ

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: PALM BAY P.D.

Serial Number: 80-001265

Time of Inspection: 11:35

Date of Inspection: 05/23/2016

Software: 8100.27

Check or Pass	YES	NO
Date and/or Flag Adjusted		No
Pre-Inspection Check (Pre-Inspection): OK	Yes	
Pre-Inspection Alcohol Test: 0.000	Yes	
Pre-Inspection Breath Sample Not Met	Yes	
Pre-Inspection Jeopard Test: Interferent Detect	Yes	
Post-Inspection Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 121116 Exp: 12/15/2016	0.08g/210L Test (g/210L) Lot#: 121116 Exp: 12/15/2016	0.20g/210L Test (g/210L) Lot#: 121116 Exp: 12/15/2016	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 12111636A2 Exp: 12/15/2016
0.000	0.044	0.080	0.197	0.080
0.001	0.048	0.080	0.197	0.080
0.002	0.048	0.080	0.198	0.080

Agency Inspector Signatures

Agency Inspector (Name) does not comply / with Chapter 11D 8, FAC.
I, _____, Florida Department of Law Enforcement Agency Inspector, certify that this inspection was conducted in accordance with the provisions of Chapter 11D 8, FAC.

Signature and Printed Name

Date

Dean Morgan