

1264
②

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: PALM BAY P.D.
Time of Inspection: 10:23

Date of Inspection: 08/23/2016

Serial Number: 80-001266
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201507A Exp: 07/14/2017	0.08g/210L Test (g/210L) Lot#:201601F Exp: 01/26/2018	0.20g/210L Test (g/210L) Lot#:201505A Exp: 05/12/2017	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:24515080A2 Exp: 10/05/2017
0.000	0.049	0.078	0.199	0.080
0.000	0.049	0.079	0.200	0.080
0.000	0.050	0.079	0.200	0.080

Number of Simulators Used: 5

Remarks:

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

EDWIN J LUTZ

Signature and Printed Name

08/23/2016
Date

D. Morgan #46

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT -- INTOXILYZER 8000

Agency: FLORIDA STATE POLICE

Serial Number: SD 601066

Inspector: DAVID MORGAN

Date of Inspection: 05/21/2016

Operator: ALOG, JF

Check or Test	YES	NO
Device Calibration Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Direct Read Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Temperature Control Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	


Alcohol Level	0.05g/210L Test (g/210L) Lot#: 2013010 Exp: 11/05/2016	0.08g/210L Test (g/210L) Lot#: 2013020 Exp: 02/04/2017	0.20g/210L Test (g/210L) Lot#: 2013010 Exp: 06/07/2016	0.09 g/210L Dry Gas Std Ver#: (g/210L) Lot#: 201300005 Exp: 10/01/2016
0.000	0.036	0.078	0.196	0.079
0.000	0.037	0.079	0.197	0.079
0.000	0.038	0.079	0.198	0.079

Agency of Inspection: FLORIDA STATE POLICE

Inspector:

This agency inspection certifies () that the device does not comply with Chapter 9, F.A.C.

I, David Morgan, hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I conducted this inspection in accordance with the provisions of Chapter 9, F.A.C.



Signature and Printed Name

Date

Date

David Morgan David Morgan

Agency of Inspection: FLORIDA STATE POLICE

Florida Department of Law Enforcement Alcohol Testing Program

RECEIVED

JAN 25 2016

FDLE

AGENCY INSPECTION REPORT - INTOXILYZER 8000 Alcohol Testing Program

Agency: PALM BAY P.D.
Time of Inspection: 08:39

Date of Inspection: 01/25/2016

Serial Number: 80-001266
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: _____ Exp: _____	0.08g/210L Test (g/210L) Lot#: _____ Exp: _____	0.20g/210L Test (g/210L) Lot#: _____ Exp: _____	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: _____ Exp: _____


Number of Simulators Used: _____

Remarks:
BYPASS AI TO OPERATE INSTRUMENT Non-compliance: COMPLIANCE NOT DETERMINED.

Compliance Not Determined.

The above instrument complies () does not comply (X) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



JAKE L SHANAHAN

Signature and Printed Name

01/25/2016
Date

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: PALM BAY P.D.
Time of Inspection: 09:31

Date of Inspection: 01/11/2016

Serial Number: 80-001266
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: _____ Exp: _____	0.08g/210L Test (g/210L) Lot#: _____ Exp: _____	0.20g/210L Test (g/210L) Lot#: _____ Exp: _____	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: _____ Exp: _____


Number of Simulators Used: _____

Remarks:
BYPASS TO OPERATE INSTRUMENT Non-compliance: .

Compliance Not determined

The above instrument complies () does not comply (X) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.


 Signature and Printed Name

JAKE L SHANAHAN

01/11/2016
 Date