

RECEIVED

# Florida Department of Law Enforcement Alcohol Testing Program

FEB 03 2016

FDLE  
Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: MIAMI-DADE PD  
Time of Inspection: 10:18

Date of Inspection: 01/20/2016

Serial Number: 80-001276  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201411K Exp: 11/25/2016	0.08g/210L Test (g/210L) Lot#:N Exp: 02/24/2017	0.20g/210L Test (g/210L) Lot#:201408B Exp: 08/12/2016	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:19014080A3 Exp: 08/05/2016
/ 0.000	0.048	0.078	0.195	0.077
/ 0.000	0.049	0.078	0.196	0.078
/ 0.000	0.049	0.078	0.197	0.078


Number of Simulators Used: 5

Remarks:

00: Ambient Fail/CHECKED AND CLEARED AREA. *CHECKED SIM*

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



PABLO ESPINOZA

Signature and Printed Name

01/20/2016  
Date