

AMENDED

# Florida Department of Law Enforcement Alcohol Testing Program

APR 29 2016  
FDLE  
Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: JACKSONVILLE SO  
Time of Inspection: 08:42

Date of Inspection: 04/28/2016

Serial Number: 80-001278  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201411K Exp: 11/25/2016	0.08g/210L Test (g/210L) Lot#:201502G Exp: 02/24/2017	0.20g/210L Test (g/210L) Lot#:201408B Exp: 08/12/2016	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:24515080A2 Exp: 10/05/2017
0.000	0.050	0.077	0.197 / 0.198	0.082
0.000	0.050	0.077	/ 0.199	0.081
0.000	0.050	0.078	/ 0.200	0.081

Number of Simulators Used: 5

Remarks:  
CHECK FOR ~~RFI~~

Notes: CHECK FOR RFI UNK. RETEST O.K.  
RST 7326

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



ROBERT D THOMASON

Signature and Printed Name

04/28/2016  
Date

*AMENDED Inspection*

RECEIVED

# Florida Department of Law Enforcement Alcohol Testing Program

APR 01 2016

FDLE

Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: JACKSONVILLE SO  
Time of Inspection: 12:27

Date of Inspection: 03/31/2016

Serial Number: 80-001278  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201411K Exp: 11/24/2016 <i>26 2117326</i>	0.08g/210L Test (g/210L) Lot#:201502G Exp: 02/24/2017	0.20g/210L Test (g/210L) Lot#:201408B Exp: 08/12/2016	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:15615080A2 Exp: 07/05/2017
0.000	0.050	0.077	0.196	0.081
0.000	0.049	0.078	0.199	0.081
0.000	0.050	0.079	0.200	0.081

Number of Simulators Used: 5

Remarks:

*Types wrong date on .05 lot - correct date 11-25-2016*

*RS  
7326*

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

*[Signature]*

ROBERT D THOMASON

Signature and Printed Name

03/31/2016  
Date