

*H.MENUEL*

RECEIVED  
APR 29 2016  
FDLE  
Alcohol Testing Program

# Florida Department of Law Enforcement Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: JACKSONVILLE SO  
Time of Inspection: 09:40

Date of Inspection: 04/28/2016

Serial Number: 80-001279  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201411K Exp: 11/25/2016	0.08g/210L Test (g/210L) Lot#: 201502G Exp: 02/24/2017	0.20g/210L Test (g/210L) Lot#: 201408B Exp: 08/12/2016	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 24515080A2 Exp: 10/05/2017
0.000	0.047	0.073 / 0.077	0.196	0.078
0.000	0.048	0.075 / 0.077	0.198	0.079
0.000	0.048	0.076 / 0.078	0.199	0.078


Number of Simulators Used: 5

Remarks:  
REPEAT INSP FOR UNK RPI WET SIM HOSE O.K MURPHY FDLE REM I.

ADD NOTES: CHECK WITH P. MURPHY R.D.L.E  
RST 7326

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



ROBERT D THOMASON

Signature and Printed Name

04/28/2016  
Date