

# Florida Department of Law Enforcement Alcohol Testing Program

**RECEIVED**

JAN 27 2016

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

**FDLE  
Alcohol Testing Program**

Agency: NASSAU COUNTY SO  
Time of Inspection: 05:31

Date of Inspection: 01/27/2016

Serial Number: 80-001281  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) <i>Amended 1-27-16</i> Lot#: 201502G <i>TJF</i> Exp: 02/24/2017 <i>201507A TJF</i>	0.08g/210L Test (g/210L) Lot#: 201502G Exp: 02/24/2017	0.20g/210L Test (g/210L) Lot#: 201505A Exp: 05/12/2017	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 22315080A3 Exp: 09/05/2017
0.000	0.049	0.077	0.199	0.081
0.000	0.049	0.078	0.200	0.082
0.000	0.049	0.078	0.200	0.081

Number of Simulators Used: 4

Remarks:

*Amended Lot. 201507A. TJF*  
*Exp. Date 7-14-2017 TJF*

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

*Terry L Nye* TERRY L NYE  
Signature and Printed Name

01/27/2016  
Date