

Florida Department of Law Enforcement Alcohol Testing Program

1287
RECEIVED
JAN 04 2017
FDLE
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: BAKER COUNTY SO
 Time of Inspection: 12:27
 Date of Inspection: 12/29/2016
 Serial Number: 80-001287
 Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	
	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201603D Exp: 03/08/2018	0.08g/210L Test (g/210L) Lot#:201601F Exp: 01/26/2018	0.20g/210L Test (g/210L) Lot#:201604C Exp: 04/05/2018	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:24515080A Exp: 10/05/2017
0.000	0.048	0.077	0.196	0.080
0.000	0.048	0.078	0.197	0.080
0.000	0.049	0.078	0.197	0.080

Number of Simulators Used: 5

Remarks:

BAKER COUNTY
SHERIFF'S OFFICE
DEC 29 2016
RECORDS SECTION
RECEIVED

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Corey S Wilson

COREY S WILSON
Signature and Printed Name

12/29/2016
Date

1257

RECEIVED

Florida Department of Law Enforcement Alcohol Testing Program

DEC 01 2016

FDLE

AGENCY INSPECTION REPORT - INTOXILYZER 8000 Alcohol Testing Program

Agency: BAKER COUNTY SO
Time of Inspection: 11:15

Date of Inspection: 11/26/2016

Serial Number: 80-001287
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201603D Exp: 03/08/2018	0.08g/210L Test (g/210L) Lot#:201601F Exp: 01/26/2018	0.20g/210L Test (g/210L) Lot#:201604C Exp: 04/05/2018	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:24515080A1 Exp: 10/05/2017
0.000	0.049	0.079	0.198	0.082
0.000	0.049	0.079	0.198	0.081
0.000	0.049	0.079	0.198	0.081

Number of Simulators Used: 5

Remarks:

BAKER COUNTY
SHERIFF'S OFFICE

NOV 28 2016

RECORDS SECTION
RECEIVED

The above instrument complies (X) does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

COREY S WILSON

Signature and Printed Name

11/26/2016
Date

RECEIVED

NOV 07 2016

Florida Department of Law Enforcement

1287



Alcohol Testing Program

FDLE

Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: BAKER COUNTY SO
Time of Inspection: 00:57

Date of Inspection: 10/30/2016

Serial Number: 80-001287
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201603D Exp: 03/08/2018	0.08g/210L Test (g/210L) Lot#:201601F Exp: 01/26/2018	0.20g/210L Test (g/210L) Lot#:201604C Exp: 04/05/2018	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:24515080A1 Exp: 10/05/2017
0.000	0.049	0.077	0.196	0.081
0.000	0.049	0.078	0.197	0.081
0.000	0.049	0.079	0.197	0.080

Number of Simulators Used: 5

Remarks:

BAKER COUNTY
SHERIFF'S OFFICE

OCT 31 2016

RECORDS SECTION
RECEIVED

The above instrument complies (X) does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

COREY S WILSON

Signature and Printed Name

10/30/2016
Date

Florida Department of Law Enforcement Alcohol Testing Program

RECEIVED

AUG 25 2016

FDLE

AGENCY INSPECTION REPORT - INTOXILYZER 8000 Alcohol Testing Program

Agency: BAKER COUNTY SO
Time of Inspection: 03:10

Date of Inspection: 08/22/2016

Serial Number: 80-001287
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201603D Exp: 03/08/2018	0.08g/210L Test (g/210L) Lot#:201601F Exp: 01/26/2018	0.20g/210L Test (g/210L) Lot#:201604C Exp: 04/05/2018	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:24515080A1 Exp: 10/05/2017
0.000	0.048	0.078	0.197	0.082
0.000	0.048	0.078	0.198	0.081
0.000	0.047	0.079	0.199	0.081

Number of Simulators Used: 5

Remarks:


BAKER COUNTY
SHERIFF'S OFFICE

AUG 22 2016

RECORDS SECTION
RECEIVED

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



COREY S WILSON

Signature and Printed Name

08/22/2016
Date

RECEIVED

Florida Department of Law Enforcement Alcohol Testing Program

AUG 04 2016

FDLE
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: BAKER COUNTY SO
Time of Inspection: 10:11

Date of Inspection: 07/31/2016

Serial Number: 80-001287
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201507A Exp: 07/14/2017	0.08g/210L Test (g/210L) Lot#:201502G Exp: 02/24/2017	0.20g/210L Test (g/210L) Lot#:201604C Exp: 04/05/2018	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:24515080A1 Exp: 10/05/2017
0.000	0.048	0.077	0.197	0.080
0.000	0.048	0.077	0.198	0.079
0.000	0.049	0.077	0.198	0.079

Number of Simulators Used: 5

Remarks:

BAKER COUNTY
SHERIFF'S OFFICE
JUL 29 2016
RECORDS SECTION
RECEIVED

The above instrument complies (X) does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

COREY S WILSON

Signature and Printed Name

07/31/2016
Date

RECEIVED

Florida Department of Law Enforcement Alcohol Testing Program

MAY 26 2016

FDLE
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: BAKER COUNTY SO
Time of Inspection: 02:28

Date of Inspection: 05/20/2016

Serial Number: 80-001287
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201507A Exp: 07/14/2017	0.08g/210L Test (g/210L) Lot#:201502G Exp: 02/24/2017	0.20g/210L Test (g/210L) Lot#:201505A Exp: 05/12/2017	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:24515080A1 Exp: 10/05/2017
0.000	0.049	0.077	0.200	0.081
0.000	0.049	0.078	0.201	0.082
0.000	0.050	0.078	0.202	0.081

Number of Simulators Used: 5

Remarks:

BAKER COUNTY
SHERIFF'S OFFICE
MAY 23 2016
RECORDS SECTION
RECEIVED

The above instrument complies (X) does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Corey S Wilson 265
Signature and Printed Name COREY S WILSON

05/20/2016
Date

RECEIVED

Florida Department of Law Enforcement MAY 06 2016

Alcohol Testing Program

FDLE Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: BAKER COUNTY SO
Time of Inspection: 01:50

Date of Inspection: 04/30/2016

Serial Number: 80-001237
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201507A Exp: 07/14/2017	0.08g/210L Test (g/210L) Lot#:201502G Exp: 02/24/2017	0.20g/210L Test (g/210L) Lot#:201505A Exp: 05/12/2017	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:24515080A1 Exp: 10/05/2017
0.000	0.049	0.077	0.199	0.082
0.000	0.049	0.078	0.200	0.082
0.000	0.049	0.078	0.201	0.082

Number of Simulators Used: 5

Remarks:

BAKER COUNTY SHERIFF'S OFFICE
MAY 02 2016
RECORDS SECTION RECEIVED

The above instrument complies (X) does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

COREY S WILSON

Signature and Printed Name

04/30/2016
Date

Florida Department of Law Enforcement RECEIVED

Alcohol Testing Program

MAY 06 2016

AGENCY INSPECTION REPORT - INTOXILYZER 8000 FDLE Alcohol Testing Program

Agency: BAKER COUNTY SO
 Time of Inspection: 05:04

Date of Inspection: 03/29/2016

Serial Number: 80-001287
 Software: 8100-27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201507A Exp: 07/14/2017	0.08g/210L Test (g/210L) Lot#:201502G Exp: 02/24/2017	0.20g/210L Test (g/210L) Lot#:201505A Exp: 05/12/2017	0.08 g/210L Dry Gas Std (g/210L) Lot#: 201501A1 Exp: 05/2017
0.000	0.049	0.078	0.199	0.080
0.000	0.049	0.078	0.200	0.080
0.000	0.050	0.079	0.201	0.080

Number of Simulators Used: 5

Remarks:

BAKER COUNTY
 SHERIFF'S OFFICE
 MAY 02 2016
 RECORDS SECTION
 RECEIVED

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Corey S Wilson COREY S WILSON
 Signature and Printed Name

03/29/2016
 Date

88-001287

RECEIVED

FEB 22 2016

Florida Department of Law Enforcement Alcohol Testing Program

FDLE
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: BAKER COUNTY SO
Time of Inspection: 01:16

Date of Inspection: 02/05/2016

Serial Number: 80-001287
Software: 8100

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201507A Exp: 07/14/2017	0.08g/210L Test (g/210L) Lot#:201502G Exp: 02/24/2017	0.20g/210L Test (g/210L) Lot#:201505A Exp: 05/12/2017	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:0281-08023- Exp: 02/02/2016
0.000	0.050	0.078	0.202	0.082
0.000	0.049	0.077	0.203	0.082
0.000	0.049	0.080	0.204	0.082

Number of Simulators Used: 5


Remarks:

BAKER COUNTY
SHERIFF'S OFFICE
FEB 16 2016

RECORDS SECTION
RECEIVED

The above instrument complies (X) does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



Signature and Printed Name COREY S WILSON

02/05/2016
Date

RECEIVED

Florida Department of Law Enforcement Alcohol Testing Program

FEB 02 2016

FDLE
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: BAKER COUNTY SO
Time of Inspection: 05:49

Date of Inspection: 01/28/2016

Serial Number: 80-001287
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201411K Exp: 11/25/2016	0.08g/210L Test (g/210L) Lot#:201502G Exp: 02/24/2017	0.20g/210L Test (g/210L) Lot#:201505A Exp: 05/12/2017	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:02814080A3 Exp: 03/01/2016
0.000	0.049	0.077	0.200	0.081
0.000	0.049	0.078	0.201	0.081
0.000	0.049	0.079	0.201	0.081

Number of Simulators Used: 5

Remarks:

BAKER COUNTY
SHERIFF'S OFFICE

JAN 28 2016

RECORDS SECTION
RECEIVED

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Corey S Wilson 263 COREY S WILSON
Signature and Printed Name

01/28/2016
Date