

RECEIVED

NOV 17 2016

FDLE

# Florida Department of Law Enforcement Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000 Alcohol Testing Program

Agency: SPRINGFIELD PD  
Time of Inspection: 15:08

Date of Inspection: 11/17/2016

Serial Number: 80-001313  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201507A Exp: 07/14/2017	0.08g/210L Test (g/210L) <del>201601P</del> Lot#: <del>201604C</del> Exp: <del>04/05/2018</del> <b>01/26/2018</b>	0.20g/210L Test (g/210L) Lot#: 201604C Exp: 04/05/2018	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 09216080A2 Exp: 05/05/2018
0.000	0.046	0.074 / 0.075	0.193	0.081
0.000	0.047	0.075 / 0.076	0.194	0.081
0.000	0.047	0.076 / 0.077	0.194	0.081

Number of Simulators Used: 4

Remarks:  
08: Control Outside Tolerance.

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

 \_\_\_\_\_  
Signature and Printed Name MIKE D KING

11/17/2016  
Date