

1320  
48

RECEIVED

# Florida Department of Law Enforcement Alcohol Testing Program

DEC 09 2016

FDLE

Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: OKBEECHOBEE CO SO  
Time of Inspection: 10:31

Date of Inspection: 12/09/2016

Serial Number: 80-001320  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201507A Exp: 07/14/2017	0.08g/210L Test (g/210L) Lot#:201601F Exp: 01/26/2018	0.20g/210L Test (g/210L) Lot#:201604C Exp: 04/05/2018	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:1686121 Exp: 05/21/2017
0.000	0.049	0.080	0.201	0.079
0.000	0.049	0.080	0.201	0.080
0.000	0.049	0.080	0.203	0.079

Number of Simulators Used: 4

Remarks:

RANGE EXCEEDED.

During the alcohol free / mouth alcohol test, too much alcohol was introduced into the instrument. The test was repeated and was successful. (M)

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

MARK W GARLAND

Signature and Printed Name

12/09/2016  
Date

RECEIVED

Florida Department of Law Enforcement  
Alcohol Testing Program

OCT 18 2016

FDLE  
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: OKEECHOBEE CO SO  
Time of Inspection: 10:50

Date of Inspection: 10/18/2016

Serial Number: 80-001320  
Software: 8100.27


Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201411K Exp: 11/25/2016	0.08g/210L Test (g/210L) Lot#:201601F Exp: 01/26/2018	0.20g/210L Test (g/210L) Lot#:201505A Exp: 05/12/2017	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:1686121 Exp: 05/21/2017
0.000 / 0.000	0.048	0.079	0.202	0.079
RFI / 0.000	0.049	0.079	0.203	0.079
/ 0.000	0.049	0.079	0.204	0.079

Number of Simulators Used: 4


Remarks:

00: RFI Detect.

During the alcohol free test, a deputy was outside of the room with a handheld radio that caused a RFI. The test was repeated successfully. 

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



MARK W GARLAND

Signature and Printed Name

10/18/2016  
Date

SP

1320

RECEIVED

# Florida Department of Law Enforcement Alcohol Testing Program

MAY 03 2016

FDLE

Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: OKEECHOBEE CO SO  
Time of Inspection: 13:06

Date of Inspection: 05/03/2016

Serial Number: 80-001320  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201411K Exp: 11/25/2016	0.08g/210L Test (g/210L) Lot#:201502G Exp: 02/24/2017	0.20g/210L Test (g/210L) Lot#:201505A Exp: 05/12/2017	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:1686121 Exp: 05/21/2017
0.000	0.048	0.079	0.203	0.079
0.000	0.049	0.079	0.204	0.079
0.000	0.049	0.079	0.203	0.079

Number of Simulators Used: 4

Remarks:

A F / M A:

During the mouth alcohol portion of the test, not a sufficient amount of mouth alcohol solution was introduced into the instrument. I repeated the test introducing a sufficient amount of mouth alcohol into the instrument and it was successful. *(Signature)*

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

*Mark W. Garland*

MARK W GARLAND

Signature and Printed Name

05/03/2016  
Date