

Florida Department of Law Enforcement RECEIVED

Alcohol Testing Program (PJM) BK MAR 03 2016

AGENCY INSPECTION REPORT - INTOXILYZER 8000 FDLE Alcohol Testing Program

Agency: ORANGE COUNTY S.O.
Time of Inspection: 08:52

Date of Inspection: 03/03/2016

Serial Number: 80-001416
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: Exp:	0.08g/210L Test (g/210L) Lot#: Exp:	0.20g/210L Test (g/210L) Lot#: Exp:	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: Exp:

Number of Simulators Used: _____


Remarks:

BYPASS TO OPERATE INSTRUMENT Non-compliance: . . .

Compliance not determined

The above instrument complies () does not comply (X) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.


 Signature and Printed Name JAKE L SHANAHAN

03/03/2016
 Date